

**DEPARTMENT OF PEDIATRICS
ALL INDIA INSTITUTE OF MEDICAL SCIENCES**

19 November 2010

To

**Subject: Short rate enquiry for the item to be purchased under the head 'Plan' year 2010-11.
Our Reference: Short Rate Enquiry no. 20/Paed/2010-2011**

Dear M/s.

A short rate enquiry is hereby floated/invited for the following item required by Department of Pediatrics, AIIMS, New Delhi.

| | |
|------------------|-----------------|
| Item Name | Minspin |
| Quantity | 02 (Two) |

Technical Specification

Main Unit:

- Compact, light weight. mini centrifuge (non refrigerated) for speed up to 14,500 rpm
- G value should be more than 14,100
- RPM/ RCF settings adjustable as required
- Rotor imbalance detection
- Digital display of speed
- Should confirm to CE IVD & IEC 1010-2-020 safety regulations.
- Certifications should be substantiated with documentary proof
- Timer from 15 sec to 99 minutes with continuous function
- Separate short spin button
- Metal rotor housing.
- Brushless motor
- Acceleration to reach max speed in <15 sec
- Braking time max. speed <15 sec.
- Power: 230 V/ 50-60 Hz

Rotors & Adapters:

- Fixed angle rotor should be at 45°
- RPM/ RCF as mentioned above
- Rotors and lids PTFE coated
- Aerosol tight
- Stain Less Steel Rotor lid with knurled nut
- Lid latch motorized
- Auto Opening of lid after the run.
- Autoclavable rotors & lids (121° C, 20 mins)

Specification continue on next page

Specification Minspin continue form previous page

Two Rotors:

- **1st Rotor** : Standard Rotor with lid
- Capacity: 12 x 2 ml & 1.5 ml micro centrifuge tubes
- **Adapters for 1 st Rotor:** Adaptors for 0.2, 0.4, 0.5 ml & 0.6 ml
Microtainer tubes = 12 nos. each
- **2nd Rotor:** PCR Strip Rotor with lid
- Capacity: 2 x 8 PCR tube strip

Conditions:

1. Features in the quotations should be substantiated with proper Principal Company Catalogue.
2. Should operate under a Quality Management System which complies with the requirements of ISO 9001:2008 for design, manufacture and services
3. Copies of all certifications e.g. Quality Standard certificate, Propriety Item/parts, Patent of parts/ technology, Principal company/Authorized Distributorship should be attached with the quotations.
4. In-House Service Engineers from principal company/ or authorised agents should be available on one-day notice basis in case of emergency.
5. Only Principal Companies or Authorized Distributors from Principal companies should quote . Quotations from non - authorized distributors will not be entertained

Terms & Conditions:

- The quotations should be submitted to the Professor & Head, Department of Pediatrics, Room No. 3058, AIIMS, New Delhi-29 and sent by post/per bearer **by 04 December 2010 up to 1.00 P.M.** Quotation should be sealed in an envelope and the name of the item clearly written on top of the envelop.
- **The firm should submit two bids in a separate envelope and both bids packed in a single envelope mentioning name of the quoted item.**

First bid 'TECHNICAL BID' should contain information of quoted equipment including all supporting documents

Second bid 'PRICE BID' should contain cost of the offered equipment. A separate list of the consumables accessories/ reagents/ spare parts etc. required to operate the equipment should be is also enclosed.

Please ensure that the cost is not disclosed in the technical bid, or else the quotation will be treated as cancelled.

- The make of the article offered should invariably be quoted. Quotation should be typed written in ink. All overwriting and erased entries will be deleted from the quotation. The rates should be valid for at least FIVE MONTHS.
- VAT/Sales tax or any other kind of tax(s) must be separately mentioned against each item. In case no sales tax is chargeable, prices must be quoted as NET.
- **Warranty/Guarantee:** 2 years warranty with spares and 3 years without spares.
- The firm should be able to give demonstration of the item if required within 1 week of requisition.
- The firm must supply the item within 45 days after confirmation of order.
- The successful bidder has to submit performances bank guarantee of 10% of the order value valid upto warranty period.
- The payment will be made electronically viz RTGS/NEFT against delivery and satisfactory installation of the item. The following information should be also mentioned in the quotation.
 1. Name of the beneficiary:
 2. Account No. of the beneficiary:
 3. IFCS code of the bank/branch:
- The quantity can be increased or decreased to any extent depending upon the actual requirement.
- The Professor & Head of the Department reserves the right to cancel/reject full or any part of the rate enquiry which generally do not fulfil the conditions stipulated in the rate enquiry.
- Note: The specifications of the items are also available at the Institute website www.aiims.edu & www.aiims.ac.in

Dr. V K Paul
Professor & Head

Dr. Ramesh Agarwal
Assistant Professor &
Faculty in-charge Deptt. Stores

Following companies have been requested to participate in limited rate enquiry for the item 'Minspin'.

Reference : Short Rate Enquiry no. 20/Paed/2010-2011

- 1. M/S Micro Devices Metrohm Limited
2/15, 3A, 3rd Floor,
West Patel Nagar,
New Delhi – 110 008**

- 2. M/S Del -Tech Instruments Pvt. Ltd.
G-58, First Floor, Vikas Marg,
Laxmi Nagar,
Delhi - 110092**

- 3. M/S Eppendorf India Ltd.
24, First Floor, Community Center,
East of Kailash,
New Delhi – 110065**

- 4. M/S Hysel India Pvt. Ltd
1405, Chranjiv Tower,
43, Nehru Place
New Delhi- 110 019**

- 5. M/s Sripad Associates
E-39,Vikas Marg
Laxmi Nagar
Delhi-110032**

- 6. M/s Nu Tek Instruments Pvt Ltd
208, Vardhman Master Plaza,
DDA, LSC, Ghazipur,
Delhi – 110 096**