

Bid Number: GEM/2025/B/5980160

Published

Bid/RA Section	Seller Query/Representation	Buyer Response	Submitted On
Terms and Conditions	As added The entire system should be from same OEM" To be deleted as there is MII content also asked	MAY BE READ AS UNDER LED OT Light domes, Mounting system and Camera should be from same OEM, Medical Grade recording system. And monitor should be from same OEM. Maximum of two OEMs allowed to quoted.	12-07-2025 04:01 PM
	MIl products must have CDSCO MD5 license	Consist of one main and One Satellite Light heads in Round Shape, Manufactured by a CE/US FDA/ Equivalent Indian standard like CDSCO etc. approved and able to provide good post sale maintenance support. MII products must have CDSCO MD5 license	12-07-2025 04:01 PM
Bid Eligibility Criteria	ITEM CATEGORY : CEILING OT LIGHT (V2) (Q2) AMENDMENT REQUIRED : We request you to please Change the tender category from V2 to Universal Category on the GeM portal, which will enable broader participation from vendors, including those current... Read more	Remain Same	12-07-2025 06:08 PM
Product/Service specification	Technical Specifications: Under Point No. 4 – Laser Pointer in the centre to see the centre of light and centre of camera. Amendment required : It should be optional to have Laser Pointer in the centre to see the centre of light and centre o... Read more	Remain Same	12-07-2025 06:08 PM
	Technical Specifications: Point No. 5 - "LED life of 60000 hrs or more with LED timer available to measure the number of hours". Amendment required : It should be optional to have LED timer available to measure the number of hours	Remain Same	12-07-2025 06:08 PM