

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-110029
STORE SECTION (DO)**

Ref. No. XX-71/SO(DO)/Gastro/2025-26/M&E


Dated: 09.09.2025

CORRIGENDUM

It is informed that the tender bearing no. XX-71/SO (DO)/Gastro/2025-26/M&E and CPP tender I.D. No. 2025_AIMSD_872065_1, was published through CPP portal has revised technical specification & dates of tender has been extended, as per TSEC recommendation and detail given below:

Procurement of Digital Subtraction Angiography System one unit on buy-back basis.			
Sl. No.	DETAIL	EXISTING	AMENDED/EXTENDED
1.	Bid submission END Date	09.09.2025	24.09.2025
2.	Bid Opening Date	10.09.2025	25.09.2025
3.	Technical Specification	Enclosed revised technical specification. Annexure-I	

All other terms & condition of the tender are same.



Sr. STORE OFFICER (DO)

ANNEXURE-I

After the discussion the TSEC members revised the points of specification as follows :-

Current Specification		Revised Specification (modifications suggested in TSEC pre-bid meeting with the company representatives)
	Mechanical System: C-Arms and Table	
8.	The table should have motorized axis tilt capacity (head-down and head - up). The tilt facility should be at least +/-15 degrees. The system should have facility for fitting the table and maintaining the C-arm projection during a longitudinal tilt.	The table should have motorized axis tilt capacity (head-down and head - up). The tilt facility should be at least +/-15 degrees. The system should have facility for tilting the table and maintaining the C-arm projection during a longitudinal tilt.
B.	X-Ray Tube:-	
3.	Anode heat storage capacity should be 3.3 MHU or more.	Anode heat storage capacity should be 3.0 MHU or more.
M.	The layout of the room should be decided by the vendor keeping in mind the requirement of endoscopy procedure & DSA.	The layout of the room should be decided by the vendor keeping in mind the requirement of endoscopy procedure & DSA. Company should inspect the site, If required, lead lining of the wall should be done. Flooring should be of granite flooring with conductive PVC flooring.
H.	Miscellaneous	
7.	Onsite visits of application support are to be guaranteed for one week after installation and as & when required	Onsite visits of application support are to be guaranteed for one week after installation and as & when required during warranty period.
	Miscellaneous	
9	Addition	Should provide Overhead cabinet: 03 No. (1200x600x1100mm) with shelves and glass shutters, Storage Cabinet: 02 No. (1200x600x2100mm) with legs.



Organisation Chain :	All India Institute of Medical Science-New Delhi Store Main Hospital - AIIMS New Delhi
Tender ID :	2025_AIMSD_872065_1
Tender Ref No :	XX-71-SODO-Gastro-2025-26-ME
Tender Title :	Purchase of Digital Subtraction Angiography System one unit on buy-back basis
Corrigendum Type :	Date

Corrigendum:1

Corrigendum Title	Corrigendum Description	Published Date	Document Name	Doc Size(in KB)
Corrigendum of bid dates extended	Corrigendum	09-Sep-2025 11:44 AM	Dateextended.pdf	266.06

Critical Dates

Publish Date	05-Aug-2025 06:00 PM	Bid Opening Date	25-Sep-2025 03:00 PM
Document Download/Sale Start Date	05-Aug-2025 06:00 PM	Document Download/Sale End Date	24-Sep-2025 03:00 PM
Clarification Start Date	05-Aug-2025 06:00 PM	Clarification End Date	12-Aug-2025 04:00 PM
Bid Submission Start Date	26-Aug-2025 04:00 PM	Bid Submission End Date	24-Sep-2025 03:00 PM

Details Before Corrigendum

Critical Dates

Publish Date	05-Aug-2025 06:00 PM	Bid Opening Date	10-Sep-2025 03:00 PM
Document Download/Sale Start Date	05-Aug-2025 06:00 PM	Document Download/Sale End Date	09-Sep-2025 03:00 PM
Clarification Start Date	05-Aug-2025 06:00 PM	Clarification End Date	12-Aug-2025 04:00 PM
Bid Submission Start Date	26-Aug-2025 04:00 PM	Bid Submission End Date	09-Sep-2025 03:00 PM



Organisation Chain :	All India Institute of Medical Science-New Delhi Store Main Hospital - AIIMS New Delhi
Tender ID :	2025_AIMSD_872065_1
Tender Ref No :	XX-71-SODO-Gastro-2025-26-ME
Tender Title :	Purchase of Digital Subtraction Angiography System one unit on buy-back basis
Corrigendum Type :	Technical Bid

Corrigendum Document Details

Corr.No.	Corrigendum Title	Corrigendum Description	Published Date	Document Name	Doc Size(in KB)
1	Corrigendum of Same amendment of tech. specification	Corrigendum	09-Sep-2025 11:51 AM	Revisedspecification.pdf 	616.25