



डॉ. राजेंद्र प्रसाद नेत्र विज्ञान केंद्र
Dr. Rajendra Prasad Centre for Ophthalmic Sciences
अखिल भारतीय आयुर्विज्ञान संस्थान , नई दिल्ली
All India Institute of Medical Sciences, New Delhi

Date: 12/10/2026

CORRIGENDUM

Sub: **Procurement of Three (03) units of “Optical Coherence Tomography (including console and catheter)” at Dr. R.P.C.**

Global Tender Ref. No.: **08/CPPP-GT/RPC/SSK/2026-27, dt. 12/09/26 (CPPP ID:2026_AIMSD_925982_1)**

It is informed that a tender of “**Optical Coherence Tomography (including console and catheter) - 03 Nos.” vide No. 08/CPPP-GT/RPC/SSK/2026-27, dt. 11/09/26 (CPPP ID:2026_AIMSD_925982_1)** was published through CPP portal. In view of representations received from prospective bidders, tendered specifications of aforesaid equipment had been amended, as per detail given below:

S. No.	Tendered Specification serial No.	Existing Published Specifications	Ammended Specifications
1	Point No. 04	The axial resolution should be ≤ 5 microns and transverse resolution (optical) should be ≤ 10 microns.	The axial resolution should be ≤ 10 microns and transverse resolution (optical) should be ≤ 20 microns.

All prospective manufactures/suppliers of the aforesaid equipment are hereby informed to take note of above, amendment only, all other tendered specifications, terms and conditions shall remain unchanged.

This is for information and records only.

STORES OFFICER
(Dr. R.P. Centre)