

## Bid Corrigendum

GEM/2026/B/7538538-C5

Following terms and conditions supersede all existing "Buyer added Bid Specific Terms and conditions" given in the bid document or any previous corrigendum. Prospective bidders are advised to bid as per following Terms and Conditions:

### Buyer Added Bid Specific Additional Terms and Conditions

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. Make in india specific authorisation certificate needs to be enclosed.
4. Actual delivery (and Installation & Commissioning (if covered in scope of supply)) is to be done at following address  
Department of Otorhinolaryngology and Head-Neck Surgery (ENT)  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES,  
ANSARI NAGAR,  
NEW DELHI-110 029.  
Free Delivery at Consignee's Site(s)
5. **Bidder financial standing:** The bidder should not be under liquidation, court receivership or similar proceedings, should not be bankrupt. Bidder to upload undertaking to this effect with bid.
6. Bidders are advised to check applicable GST on their own before quoting. Buyer will not take any responsibility in this regards. GST reimbursement will be as per actuals or as per applicable rates (whichever is lower), subject to the maximum of quoted GST %.
7. Malicious Code Certificate:  
  
The seller should upload following certificate in the bid:-  
(a) This is to certify that the Hardware and the Software being offered, as part of the contract, does not contain Embedded Malicious code that would activate procedures to :-  
(i) Inhibit the desires and designed function of the equipment.  
(ii) Cause physical damage to the user or equipment during the exploitation.  
(iii) Tap information resident or transient in the equipment/network.  
(b) The firm will be considered to be in breach of the procurement contract, in case physical damage, loss of information or infringements related to copyright and Intellectual Property Right (IPRs) are caused due to activation of any such malicious code in embedded software.
8. **Manufacturer Authorization:**Wherever Authorised Distributors/service providers are submitting the bid, Authorisation Form /Certificate with OEM/Original Service Provider details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid
9. **OPTION CLAUSE:** The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 50 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 50% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be (Increased quantity ÷ Original quantity) × Original delivery period (in days), subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.

10. Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
11. Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
12. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.
13. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
14. Bidder / OEM has to give an undertaking that after expiry of warranty period, it will provide Comprehensive Maintenance Service for next 2 years for the offered products at the rate not more than 5 % of contract price per annum. Buyer reserves the right to enter into a CMC agreement with the Successful Bidder / OEM after expiry of the Warranty period at above mentioned rate and the payment for the CMC charges would be made Annually after rendering of the CMC Services of the relevant CMC period. Performance Security of the successful bidder shall be forfeited if it fails to accept the CMC contract when called upon by the buyer. CMC would include cost of All components including spares/accessories/labor and third party item (if any) required for installation and fully functional of equipment.  
(Upload the undertaking). The original Performance Security of contract will be returned only after submission and verification of AMC Performance Security for 3% of total CMC value valid up to CMC period plus 2 months (if there is no other claim).
15. Warranty period of the supplied products shall be as given in specifications from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid.
16. Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope  
All components including spares/accessories/labor and third party item (if any) required for installation and fully functional of equipment.  
(Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)
17. Buyer uploaded ATC document [Click here to view the file](#).
18. Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
19. ISO 9001: The bidder or the OEM of the offered products must have ISO 9001 certification.
20. The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
21. Material Test Certificate Should Be Sent Along with The Supply. The Material Will Be Checked by Buyer's Lab & the Results of the Lab will be the Sole Criteria for Acceptance of the Item.
22. Bidders can also submit the EMD with Account Payee Demand Draft in favour of  
AIIMS MAIN GRANT  
payable at  
NEW DELHI

Bidder has to upload scanned copy / proof of the DD along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date / Bid Opening date.

23. Bidders can also submit the EMD with Fixed Deposit Receipt made out or pledged in the name of A/C AIIMS MAIN GRANT  
. The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of EMD, the FDR will be released in the favour of the bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Bidder has to upload scanned copy/ proof of the FDR along with bid and has to ensure delivery of hardcopy to the Buyer within 5 days of Bid End date/ Bid Opening date
24. Successful Bidder can submit the Performance Security in the form of Account Payee Demand Draft also (besides PBG which is allowed as per GeM GTC). DD should be made in favour of AIIMS MAIN GRANT payable at NEW DELHI  
. After award of contract, Successful Bidder can upload scanned copy of the DD in place of PBG and has to ensure delivery of hard copy to the original DD to the Buyer within 15 days of award of contract.
25. Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of AIIMS MAIN GRANT A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.
26. Bidder's offer is liable to be rejected if they dont upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
27. AVAILABILITY OF OFFICE OF SERVICE PROVIDER: An office of the Service Provider must be located in the state of Consignee. DOCUMENTARY EVIDENCE TO BE SUBMITTED.
28. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
29. Buyer Added text based ATC clauses

**Bidders are advised kindly submit their bid as per revised Technical Specification details as under:-**

## **REVISED TECHNICAL SPECIFICATION**

### **(ANNEXURE-1)**

<b>Item Name: Plasma Ablation System</b>	
<b>Qty: 03 Nos</b>	
<b>Warranty Guarantee: 2 Years Warranty+ 8 Years C AMC</b>	
<b>A</b>	<b>Main Unit- 01 No</b>

1	Ablation System Should be based on low temperature Bipolar radio frequency technology.
2	The output voltage setting should be controlled and regulated from setting button on the generator.
3	The console should withstand power interruption while unplugging and re-plugging.
4	Console should have acoustic feedback signal with adjustable volume.
5	The RF probe should be able to produce a uniform production of plasma
6	Thermal Penetration should not be more than 160-180 Micron
7	USFDA/European CE/ISO/ any Indian equivalent standard
8	All the consumable and consumable attached to device must be supplied by same manufacturer.
9	The supplier should provide alternative console when the machine is taken up for repair till the repair of the device
10	Controlled Plasma ablation of tissues based on low temperature frequency technology in electrolytic solution like normal saline. The ablation setup ranged from 0 to 9 and coagulation ranged from 0 to 5
11	Operating Temperature must be in range between 40 to 70 degree
12	The generator should have integrated automated saline pump for continuous irrigation of the surgical site.
13	The output voltage setting should be controlled by regulation.

14	The generator should have facility to adjust Ablation and Coagulation with different setting.
15	Probes should have integrated cable.
16	Electrodes should be made of tungsten for enhanced probe life
17	Probes should have a malleable shaft.
18	Suction probes should have integrated IV tubing
19	Probes should have gravity defying saline delivery system to ensure saline reaches tip irrespective of probes orientation.
20	Trolley-01 Nos
21	Disinfecting Tray to be provided for the wands-02 Nos
<b>B</b>	<b>Foot Pedal Switch-01 No</b>
1	Reusable foot pedal should be waterproof/if not waterproof then plastic cover to be quoted separately.
2	There should be facility to adjust Radiofrequency Ablation system and Coagulation with different settings as per procedure demand.
<b>C</b>	<b>Consumable for 01 year</b>
1.	Adeno Tonsillectomy wands with malleable shaft with suction, irrigation, coagulation and ablation, Shaft Length(13-14cm) Diameter 5-6mm <b>-20 Nos</b>
2.	Turbinate Wands for sub-mucosal turbinate resection Length 10-12cm Diameter 1.8-2.0mm <b>-05 Nos</b>

3.	Malleable Laryngeal Wands should have atleast 16-17cm working length <b>-05 Nos</b>
4.	Malleable Micro laryngeal Wands, Shaft Length(19-20cm) Diameter(2.7-3.0mm) <b>-05 Nos</b>
5.	Adenoidectomy Wand Length 13-14cm. Diameter 2.8-3.0 mm X 5.3-5.5mm with 1.8-2.0mm suction <b>-05 Nos</b>
6.	Soft Palate Wand Diameter 1.6-1.9mm Working Length 10-11cm <b>-05 nos</b>
<b>D</b>	<b>Portable EndoScopy Unit-02</b>
1.	Should have 4 in 1 (Camera + LED Light Source + Monitor + HD Recorder)
2.	Should have LCD screen size 17 -18 inch
3.	Should have resolution 1900-1920 X 1000-1080P in displaying, recording/Storing and replaying.
4.	Should have recording format MP4/MPEG4 or latest
5.	Should have storage media USB flash disk, Pendrive / HDD
6.	Light Weight Camera Head with Cable atleast 2.5metre or more in length
7.	Should have LED Brightness of 12-16 LUX
8.	Should have LED life expectancy 20 thousand hours.
9.	Should have colour temperature 6000-6500 K
10.	Supplied with Hard Storage case
<b>E</b>	<b>Instruments</b>

1.	Zero Degree Endoscope of length 18-19cm with universal Adaptor and metal Storage Case
2.	Thirty Degree Endoscope of length 18-19cm with universal Adaptor and metal Storage Case
3.	Rotatable Back Biting-01
4.	Giraffe Forceps for Frontal Sinus side biting-01
5.	Giraffe Forceps for Frontal Sinus AP biting-01
6.	Malleable Suction-01
7.	Straight Nasal Suction -01
8.	Curve Nasal suction 40 Degree-01
9.	Curve Nasal suction 60 Degree-01
10.	Sickle Knife-01
11.	Blakesley Forceps Zero Degree-01
12.	Blakesley Forceps Forty Five Degree-01
13.	Blakesley Forceps Ninety Degree-01
14	Thru Cut Forceps Zero Degree-01
.15.	Thru Cut Forceps Forty Five Degree-01
16	Thru Cut Forceps Ninety Degree-01
17	Double End Freer 1-2mm and 3-5 mm Tip, Length 18-20 cm -01
18	Double End Freer 1-2mm and 3-5 mm Tip, Length 20-25 cm -01

19	Rotatable Kerrison Straight 3mm bite-01
20	Rotatable Kerrison Straight 1mm bite-01
21	Neurosurgery Endoscopic Scissor Straight-01
22	Neurosurgery Endoscopic Scissor Curved-01
23	Neurosurgery Long Blakesley Forceps-01
24	Nasal Packing Forceps-01
25	Tilly Hackle Forceps-01
26	BP Handle-01
27	Osteotome 1mm-01
28	Osteotome 2mm-01
29	Osteotome 5mm-01
30	Hammer-01
31	Sterilization box with silicon matt for storing Instruments-01
32	Ball Probe Double End-01
<b>F</b>	<b>Physical Demo of Machine with all wands should be provided at our premises.</b>
<b>G</b>	<b>Warranty and CAMC should be given to everything except the consumable mentioned at 'C and E'</b>
<b>H</b>	<b>All the consumable/Accessory should be maintained and supplied by same vendor.</b>

<b>I</b>	<b>Company should quote the price of all Consumables/Accessory and price of all consumable/Accessory should be freeze for upcoming 10 Years</b>
<b>J</b>	<b>The Principle/Supplier should maintain and support the system for upcoming 10 years (warranty + CAMC) if the existing system or any other part is obsolete or unrepairable then principle/supplier upgrade/Replace it without any additional cost.</b>
<b>K</b>	<b>System should be fully functional with all necessary software and hardware.</b>
<b>F</b>	<b>If the Maintenance contractor may have to take the goods or some components of the goods to his factory for repair on such occasion before handing the overs or component the contractor gives loaner to the department.</b>
<b>G</b>	<b>After sales services center should be available at the city of institution on 24 (hrs) X7 (days) X 365 (Days) basis. Complaints should be attended properly, maximum within 8 hrs. The service should be provided directly by bidder/Indian Agent. Undertaking by the Principals in the "Manufacturer Authorization Form" that the spares for the equipment shall be available for at least 10 years from the date of supply of equipment.</b>
<b>H</b>	<b>On site training to Doctors/Technicians/staff is to be provided by Principal/Indian Agents for operation and maintenance of the equipment to the satisfaction of the department.</b>
<b>I</b>	<b>During the warranty period and CAMC period, desired Uptime of 95% of 365/366(Leap Year) days (24 Hrs), if downtime more than 5%, the warranty period/CAMC period will be extended by twice the time period. In addition a penalty equal off 0.25% of the total cost of equipment per day will be liveable for the excess downtime period. Complaints should be attended properly maximum within 8hrs.</b>

<b>J</b>	<b>All software upgradation should be provided free of cost during Warranty and Comprehensive Warranty Period.</b>
<b>K</b>	<b>The Specification are broad based, generalized in respect of the requirement and not suited to any particular firm/brand</b>

## Disclaimer

The Additional Terms and Conditions (ATC) have been incorporated by the Buyer after approval of their Competent Authority. The Buyer, is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any restriction arising in the bidding process due to these ATCs and including the modification of technical specifications and / or terms and conditions governing the bid. All representations / grievances pertaining to the ATC clauses shall be raised with the buyer organization directly and not with GeM. If any of the clause(s) is/are incorporated by the Buyer regarding the following, the bid & resultant contract shall be treated as null & void. Further, GeM reserves the right, at its sole discretion, to cancel the bid forthwith, without issuance of any prior notice or intimation :-

1. Publishing Custom / BOQ bids for items for which regular GeM categories are available (unless such Custom / BOQ item is bunched with the major regular product Category Item).
2. Mandating procurement of / from specific Brand / Make / Model / Manufacturer / Dealer except in case of Single Bid / Proprietary Article Certificate (PAC) Buying.
3. Inclusion of disqualification criteria related to suspension of seller / service provider, where such suspension period has already expired.
4. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
5. Publishing bids on GeM for procurement of works.
6. Procurement of Goods by creating a Service bid on GeM & vice-versa.
7. Seeking sample with bid or approval of samples during bid evaluation process. However, trial / sample, as the case may be, shall be permitted in cases where trial / sample are allowed as per approved and published procurement policy of the Buyers' controlling Ministry / Department / State / Public Sector Enterprises Headquarters. If there is any violation of trial / sample clause with regard to approved policy of the Buyers' Ministry / Department / State / Public Sector Enterprises Headquarters, then this is to be determined and redressed by the concerned Buyer Organisation only.
8. Seeking experience from specific organization / department / institute only or from foreign / export experience.
9. Creating bid for items from incorrect categories.
10. Reference of conditions published on any external site or reference to external documents/clauses.
11. Asking for any Tender fee / Bid Participation fee, as the case may be.
12. Buyer added ATC Clauses which are in contravention of clauses defined in bid detail section, including specifications, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by the applicable GeM GTC.
13. Any ATC clause in contravention with GeM GTC Clause 4 (xiii) (h) will be invalid. In case of multiple L1 bidders against a service bid, the buyer shall place the Contract by selection of a bidder amongst the L-1 bidders through a Random Algorithm executed by GeM system.
14. In a category based bid, adding additional items, through buyer added, additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogues or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller. Buyer is duty bound to reply to all such

representations and would not be allowed to open bids if he fails to reply to such representations.

\*This document shall overwrite all previous versions of Bid Specific Additional Terms and Conditions.

[This Bid is also governed by the General Terms and Conditions](#)



ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI-110029  
Stores Section (DO)

Dated:- 13.06.2026

**GEM Bid No. GEM/2026/B/7538538 dated 14.05.2026**  
Tender ref. No. XX-32/SO(DO)/ENT/GEM/2026-27/M&E

**CORREGENDUM**

It is inform that the GeM Bid No. GEM/2026/B/7538538 dated 14.05.2026 and tender reference No. XX-32/SO(DO)/ENT/GEM/2026-27/M&E has been published on GeM Portal, there is some amendments as detail given below:

Sr. No.	Detail	Amended /Extended
1.	Existing Bid End Date	13.06.2026
2.	Extended up to	04.07.2026

All other terms & condition of the tender are same

*13/6/26*

SR. STORES OFFICER (DO)



ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI-110029  
Stores Section (DO)

Dated:- 22.06.2026

GEM Bid No. GEM/2026/B/7538538 dated 14.05.2026  
Tender ref. No. XX-32/SO(DO)/ENT/GEM/2026-27/M&E

**CORREGENDUM**

It is inform that the GeM Bid No. GEM/2026/B/7538538 dated 14.05.2026 and tender reference No. XX-32/SO(DO)/ENT/GEM/2026-27/M&E has been published on GeM Portal, there is some amendments as detail given below:

Sr. No.	Detail	Existing	Amended /Extended
1.	Bid Submission End date	04.07.2026	11.07.2026
2.	Technical Specification	As Per Bid Document	Revised Technical Specification Attached <b>ANNEXURE-1</b>

All other terms & condition of the tender are same

*Handwritten signature and date: 22/6/26*

**SR. STORES OFFICER (DO)**

## REVISED TECHNICAL SPECIFICATION AFTER PRE BID MEETING (ANNEXURE-1)

<b>Item Name: Plasma Ablation System</b>	
<b>Qty: 03 Nos</b>	
<b>Warranty Guarantee: 2 Years Warranty+ 8 Years CAMC</b>	
<b>A</b>	<b>Main Unit- 01 No</b>
1	Ablation System Should be based on low temperature Bipolar radio frequency technology.
2	The output voltage setting should be controlled and regulated from setting button on the generator.
3	The console should withstand power interruption while unplugging and re-plugging.
4	Console should have acoustic feedback signal with adjustable volume.
5	The RF probe should be able to produce a uniform production of plasma
6	Thermal Penetration should not be more than 160-180 Micron
7	USFDA/European CE/ISO/ any Indian equivalent standard
8	All the consumable and consumable attached to device must be supplied by same manufacturer.
9	The supplier should provide alternative console when the machine is taken up for repair till the repair of the device
10	Controlled Plasma ablation of tissues based on low temperature frequency technology in electrolytic solution like normal saline. The ablation setup ranged from 0 to 9 and coagulation ranged from 0 to 5
11	Operating Temperature must be in range between 40 to 70 degree
12	The generator should have integrated automated saline pump for continuous irrigation of the surgical site.
13	The output voltage setting should be controlled by regulation.
14	The generator should have facility to adjust Ablation and Coagulation with different setting.
15	Probes should have integrated cable.
16	Electrodes should be made of tungsten for enhanced probe life
17	Probes should have a malleable shaft.
18	Suction probes should have integrated IV tubing
19	Probes should have gravity defying saline delivery system to ensure saline reaches tip irrespective of probes orientation.
20	Trolley-01 Nos
21	Disinfecting Tray to be provided for the wands-02 Nos
<b>B</b>	<b>Foot Pedal Switch-01 No</b>
1	Reusable foot pedal should be waterproof/if not waterproof then plastic cover to be quoted separately.

2	There should be facility to adjust Radiofrequency Ablation system and Coagulation with different settings as per procedure demand.
<b>C</b>	<b>Consumable for 01 year</b>
1.	Adeno Tonsillectomy wands with malleable shaft with suction, irrigation, coagulation and ablation, Shaft Length(13-14cm) Diameter 5-6mm -20 Nos
2.	Turbinate Wands for sub-mucosal turbinate resection Length 10-12cm Diameter 1.8-2.0mm -05 Nos
3.	Malleable Laryngeal Wands should have atleast 16-17cm working length -05 Nos
4.	Malleable Micro laryngeal Wands, Shaft Length(19-20cm) Diameter(2.7-3.0mm) -05 Nos
5.	Adenoidectomy Wand Length 13-14cm. Diameter 2.8-3.0mm X 5.3-5.5mm with 1.8-2.0mm suction-05 Nos
6.	Soft Palate Wand Diameter 1.6-1.9mm Working Length 10-11cm-05 nos
<b>D</b>	<b>Portable EndoScopy Unit-02</b>
1.	Should have 4 in 1 (Camera + LED Light Source + Monitor + HD Recorder)
2.	Should have LCD screen size 17 -18 inch
3.	Should have resolution 1900-1920 X 1000-1080P in displaying, recording/ Storing and replaying.
4.	Should have recording format MP4/MPEG4 or latest
5.	Should have storage media USB flash disk, Pendrive / HDD
6.	Light Weight Camera Head with Cable atleast 2.5 metre or more in length
7.	Should have LED Brightness of 12-16 LUX
8.	Should have LED life expectancy 20 thousand hours.
9.	Should have colour temperature 6000-6500 K
10.	Supplied with Hard Storage case
<b>E</b>	<b>Instruments</b>
1.	Zero Degree Endoscope of length 18-19cm with universal Adaptor and metal Storage Case
2.	Thirty Degree Endoscope of length 18-19cm with universal Adaptor and metal Storage Case
3.	Rotatable Back Biting-01
4.	Giraffe Forceps for Frontal Sinus side biting-01
5.	Giraffe Forceps for Frontal Sinus AP biting-01
6.	Malleable Suction-01
7.	Straight Nasal Suction -01
8.	Curve Nasal suction 40 Degree-01
9.	Curve Nasal suction 60 Degree-01
10.	Sickle Knife-01
11.	Blakesley Forceps Zero Degree-01
12.	Blakesley Forceps Forty Five Degree-01
13.	Blakesley Forceps Ninety Degree-01
14.	Thru Cut Forceps Zero Degree-01
15.	Thru Cut Forceps Forty Five Degree-01
16.	Thru Cut Forceps Ninety Degree-01

17	Double End Freer 1-2mm and 3-5 mm Tip, Length 18-20cm -01
18	Double End Freer 1-2mm and 3-5 mm Tip, Length 20-25cm -01
19	Rotatable Kerrison Straight 3mm bite-01
20	Rotatable Kerrison Straight 1mm bite-01
21	Neurosurgery Endoscopic Scissor Straight-01
22	Neurosurgery Endoscopic Scissor Curved-01
23	Neurosurgery Long Blakesley Forceps-01
24	Nasal Packing Forceps-01
25	Tilly Hackle Forceps-01
26	BP Handle-01
27	Osteotome 1mm-01
28	Osteotome 2mm-01
29	Osteotome 5mm-01
30	Hammer-01
31	Sterilization box with silicon matt for storing Instruments-01
32	Ball Probe Double End-01
F	Physical Demo of Machine with all wands should be provided at our premises.
G	Warranty and CAMC should be given to everything except the consumable mentioned at 'C and E'
H	All the consumable/Accessory should be maintained and supplied by same vendor.
I	Company should quote the price of all Consumables/Accessory and price of all consumable/Accessory should be freeze for upcoming 10 Years
J	The Principle/Supplier should maintain and support the system for upcoming 10 years (warranty + CAMC) if the existing system or any other part is obsolete or unrepairable then principle/supplier upgrade/Replace it without any additional cost.
K	System should be fully functional with all necessary software and hardware.
F	If the Maintenance contractor may have to take the goods or some components of the goods to his factory for repair on such occasion before handing the overs or component the contractor gives loaner to the department.
G	After sales services center should be available at the city of institution on 24 (hrs) X7 (days) X 365 (Days) basis. Complaints should be attended properly, maximum within 8 hrs. The service should be provided directly by bidder/Indian Agent. Undertaking by the Principals in the "Manufacturer Authorization Form" that the spares for the equipment shall be available for at least 10 years from the date of supply of equipment.
H	On site training to Doctors/Technicians/staff is to be provided by Principal/Indian Agents for operation and maintenance of the equipment to the satisfaction of the department.
I	During the warranty period and CAMC period, desired Uptime of 95% of 365/366(Leap Year) days (24 Hrs), if downtime more than 5%, the warranty period/CAMC period will be extended by twice the time period. In addition a

	penalty equal off 0.25% of the total cost of equipment per day will be liveable for the excess downtime period. Complaints should be attended properly maximum within 8hrs.
J	All software upgradation should be provided free of cost during Warranty and Comprehensive Warranty Period.
K	The Specification are broad based, generalized in respect of the requirement and not suited to any particular firm/brand



(<https://gem.gov.in/>)

[PRODUCTS](#) [SERVICES](#) [CONTENT](#)

Enter Keywords?



HOME ([HTTPS://MKP.GEM.GOV.IN/DASHBOARD](https://mkp.gem.gov.in/dashboard)) / BID FINALIZATION

BID DETAILS

TECHNICAL EVALUATION

FINANCIAL EVALUATION

EVALUATION

BID AWARDED

## 1. Bid Details

Your corrigendum has been published successfully. (<https://bidplus.gem.gov.in/buyer-bid-finalization/9327648>)

Bid Number: **GEM/2026/B/7538538** (</showbidDocument/9327648>)

**Bid Status:** Active

**Quantity:** 103

**Bid Validity (From End Date):** 180 (Days)

**Competent Authority Document:**  
View

**Bid Start Date / Time:** 14-05-2026 16:54:19

**Bid End Date / Time:** 11-07-2026 17:00:00

**Bid Opening Date / Time:** 11-07-2026 17:30:00

Consignees / Reporting Officer / Delivery Location(S)

**EMD:** Required Track EMD (<https://bidplus.gem.gov.in/Bidding/Track/Trac>)

### Buyer Details

**Name:** Naveen Kumar

**Address:** Naveen

Kumar,Naveenkr@Aiims.Gov.In,1st Floor , New Pvt. Ward, Store Section (CN Centre) , AIIMS, Ansari Nagar, New Delhi,SOUTH DELHI,DELHI,110029,India,011-26594850-

**Ministry:** Ministry Of Health And Family Welfare

**Department:** Department Of Health And Family Welfare

**Organisation:** All India Institute Of Medical Sciences (Aiims)

**Office:** Aiims, New Delhi

### Corrigendum Details

**Modified On:** 2026-06-22 17:13:39

[Download](#)

(</bidding/buyer/showcorrigendumpdf/4335666/9327648>)

**Hash Value (Algorithm - SHA256):**

fddac74dfb93d24a6d15b4a7c5d01a7e88de5e68359dd7d888d1ab5a9add6c5

**Modified On:** 2026-06-22 17:01:49

**Bid extended to** 2026-07-11 17:00:00

Bid Opening Date : 2026-07-11 17:30:00

**Modified On:** 2026-06-13  
11:45:00

Bid extended to 2026-07-04 13:00:00

Bid Opening Date : 2026-07-04 13:30:00

Cancel Bid

Extend Bid

Edit Terms

Edit Pre Bid

## WEB INFO

Terms of Use (<https://gem.gov.in/termsCondition>)

Website Policies (<https://gem.gov.in/websitePolicies>)

Document Help (<https://gem.gov.in/help>)

Site Map (<https://gem.gov.in/sitemap>)

Web Information Manager (<https://gem.gov.in/web-information-manager>)

## NEWS & EVENTS

Newsroom (<https://gem.gov.in/media>)

Gallery (<https://gem.gov.in/gallery>)

Notifications (<https://gem.gov.in/landing/index/allnews>)

CCM Schedule ([https://gem.gov.in/ccm\\_data](https://gem.gov.in/ccm_data))

Forums (<https://gem.gov.in/forum>)

## TESTIMONIALS

([HTTPS://GEM.GOV.IN/TESTIMONIAL\\_BUYERS](https://gem.gov.in/testimonial_buyers))

## TRAINING

LMS (<https://lms.gem.gov.in/>)

Training Calendar (<https://gem.gov.in/training>)

Training Module ([https://gem.gov.in/training/training\\_module](https://gem.gov.in/training/training_module))

Facilitators (<https://gem.gov.in/training/facilitators>)

Download GeM Logo (<https://gem.gov.in/resources/images/download-gem-logo.jpg>)

## ABOUT GeM

Introduction to GeM (<https://gem.gov.in/aboutus>)

Statistics (<https://gem.gov.in/statistics>)

Right to Information (<https://gem.gov.in/RTI>)

Analytics (<https://sso.gem.gov.in/ARXSSO/oauth/login>)

New on GeM (<https://gem.gov.in/new-categories>)

BRAND GeM (<https://gem.gov.in/brand-gem>)

## RESOURCES

GeM Handbook ([https://assets-bg.gem.gov.in/resources/pdf/GeM\\_handbook.pdf](https://assets-bg.gem.gov.in/resources/pdf/GeM_handbook.pdf))

OM's/Circulars  
([https://gem.gov.in/support/government\\_oms\\_circulars](https://gem.gov.in/support/government_oms_circulars))

Terms and Conditions ([https://gem.gov.in/support/terms\\_conditions](https://gem.gov.in/support/terms_conditions))

Policies/Manuals (<https://gem.gov.in/support/buyers>)

Miscellaneous (<https://gem.gov.in/support/miscellaneous>)

MoUs (<https://gem.gov.in/mou>)

## NEED HELP?

FAQs (<https://gem.gov.in/userFaqs>)

Video Guides (<https://gem.gov.in/training/videos/buyers>)

Raise-a Ticket (<https://gem.gov.in/gemtickets/create>)

Contact Us (<https://gem.gov.in/contactUs>)

Careers (<https://gem.gov.in/landing/index/careers>)