

Store Section

Dr. BRAIRCH

Date: 21.03.2017

Ref. No. F. No - **239** / IRCH/SSK-1/Surgical Oncology /Proprietary/2016-2017.


Subject: - Purchase of CDH -25A on Proprietary Basis-Inviting Comments thereon.

The request letter received from Department of Surgical Oncology Dr.BRAIRCH for the procurement of the item given below on proprietary basis from M/S Johnson & Johnson.

S.N.O	Item Description
1.	CDH 25 A (Intra Luminal Circular Stapler Curved 25mm, having large diameter knife blade offers 16.4mm lumen, with detachable anvil head , with longer staple leg length (open -5.5mm) , controlled tissue compression (closed staple height 1.0mm-2.5mm)

The user Dept. has submitted Proprietary Article Certificate in addition to one submitted by company.

The above documents are being uploaded on AIIMS WEB SITE for open information to invite objection, comments, if any, from any manufacturer regarding proprietary nature of the item within 15days form the date of issue/ uploading of the notification giving reference F. No.- 239/IRCH/SSK-1/ Surgical Oncology /Proprietary/2016-2017. The Comments should be sent to Stores Officer, Dr. BRA IRCH AIIMS on or before 07 April 2017 up to 12:00 noon, failing which it will be presumed that any other vendor /company is having no comment to offer and case will be decided purely on merits.


STORES OFFICER
(DR. BRA IRCH)

Encl: Related Documents enclosed
PAC Certificates enclosed.

2

TO WHOMSOEVER IT MAY CONCERN

Circular stapler Intra Luminal Stapler(CDH25A) having curved shaft anvil detachable, 25mm diameter head, adjustable stapler height (1.0-2.5 mm) for control tissue compression, longer staple leg length of 5.5 mm and having lumen size of 16.4mm is our proprietary item.

M/s Oriental Surgical Works
1708/9, 1st Floor Mittal Medical Market,
Bhagirath Palace
Chandni Chowk
Delhi 110 006
011-23863508, 9312211408

G. Viswanathan
Regional Sales Manager,
Ethicon Endo-Surgery,
Johnson & Johnson Medical India

Dr. S. S. A. Institute Rotary Cancer Hospital
New Delhi-110029