

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI-29.  
STORES SECTION (DO)**

Ref. No. 18/Stores (DO)/Uro/PAC/2017-18/FSC

Dated 15.07.2017

14.7.17

**Sub:- Purchase of Combined Anti neo-plastic Thermotherapy Bladder Recirculation System for Ca Bladder** for the department of Urology, (AIIMS), New Delhi-110 029, on proprietary basis Inviting comments thereon.

\*\*\*\*\*

The Institute is in the process to purchase **Combined Anti neo-plastic Thermotherapy Bladder Recirculation System for Ca Bladder** for the department of Urology, (AIIMS), New Delhi from M/s. Combat Medical Ltd., U.K. through M/s. Brainrootz Labs Pvt. Ltd., New Delhi. The proposal submitted by M/s. Brainrootz Labs Pvt. Ltd. and PAC Certifications are attached.

The above documents are being uploaded for open information to submit objections, comments, if any, from any manufacturer regarding proprietary nature of the equipment/item within issue of 15 days giving reference No. 18/Stores (DO)/Uro/PAC/2017-18/FSC. The comments should be received by office of Sr. Stores Officer (DO), Store Section (DO), Animal House Building, Near Biotechnology Building at AIIMS on or before 5.8.17 upto 12.30 p.m., failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

Yours faithfully,

  
SR. STORES OFFICER (DO) 14.7.17

**Encl: Related documents enclosed.**

(17)

ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI - 110 029

PROPRIETORY/SPECIFIC BRAND GOODS CERTIFICATE

1. Item/Type/Model No. required along with specification **Combined Antineoplastic Thermoherapy Bladder Recirculation System for Ca Bladder**
2. Is the item a spare part attachment or accessory for an existing equipment **No**
3. Name of the manufacturers/Supplier of the item proposed by the indenter. **M/s Combat Medical, U.K. / M/s Brainrootz Labs Pvt. Ltd., New Delhi**
4. Are they sole manufacturers/sold distributors of the item **Yes**
5. Is there any other item with similar/equivalent specification available in the market to meet the job requirement envisaged. **No**  
If the answer is yes, why the same can't be procured. Demanding Officer should bring out comparative functional advantage/cost effectiveness of the recommended item from these offered by other.
6. What were the efforts made to locate alternative source of supply of use other substitutes. **Internet/ No other alternative source for supply of other substitutes**
7. Why open/limited tender can't be resorted to, for locating alternative source. **These are proprietary items of M/s Combat Medical, U.K.**
8. Are the proprietary items certifying that the rates are reasonable or not. **Rates certificate from the firm is enclosed**
9. Any other justification for procuring item from single source. **Only single Manufacturer**

The above mentioned information is provided by the firm.

Signature of Indenter  
(Demanding Officer)

DR. P. N. DOGRA

Head of the Department

I certify that the item at Sr. No. 1 above is required to be procured on single tender basis as the source of supply is definitely known/the specified brand proposed was advantages in meeting our functional requirements and limited tender system could be dispensed with as they would serve no useful purpose in this particular case.

(Strike our whichever is not applicable)

*[Handwritten signatures and initials]*

DR. P. N. DOGRA  
Head of the Department  
All India Institute of Medical Sciences  
Ansari Nagar, New Delhi-29, INDIA

Da 06-05-2017

To  
The HOD  
Department of Urology  
AIIMS,  
New Delhi



**Sub:- Rate Quotation for COMBAT BRS Recirculating Fluid Warmer System & Bladder Recirculation Set**

Dear Sir,

We are thankful to you for considering our original research Medical Device for your esteemed institution.

We are pleased to offer our lowest hospital rates to your esteemed institution for the current financial year (2017-2018).  
The complete details are as under:

Ref. No.	Device Discription	Qty.	Rate	Total (Excluding VAT)
COMBRS4	COMBAT BRS Recirculating Fluid Warmer System	1	20,00,000/-	20,00,000/-
COMSETBRS04	COMBAT BRS Bladder Recirculation Set	1	15,000/-	15,000/-

The rates quoted are inclusive of customs duty, in case of any revision in Rate, the same will be intimated to your office. The supplies shall be executed through

M/s Brainrootz Labs Pvt. Ltd.  
8H, Gopala Tower,  
Rajendra Place, N. Delhi - 110008  
Contact No. +91-8447056700, +91-11-45170355

We shall look forward to your valuable supply orders.

Thanking you and assuring you of our best services always,

Yours Faithfully,

Rohit TRIPATHI  
Head - Sales & Marketing  
+91-8447056700  
Brainrootz Oncology  
Brainrootz Labs Pvt. Ltd.



DR. P.N. DOGRA  
अध्यक्ष एवं अध्यक्ष/Professor & Head  
मूत्ररोग विज्ञान विभाग/Deptt. of Urology  
अखिल भारतीय आयुर्विज्ञान संस्थान  
All India Institute of Medical Sciences  
अंसारी नगर, नई दिल्ली/Ansari Nagar, New Delhi-28, INDIA

(9)

Date: - 03/07/2017

To

The Director

All India Institute of Medical Sciences

New Delhi



Brainrootz Labs

**Terms & Conditions**

1. Validity : 180 days or upto 31 Dec 2017
2. Payment : 100% against satisfactory installation
3. Warranty : 5 years for manufacturing defects only from the date of installation
4. CMC : CMC charges for 5 years on FOB / FOR Prices.

1 <sup>st</sup> year	:	3 %
2 <sup>nd</sup> year	:	3.5 %
3 <sup>rd</sup> year	:	4%
4 <sup>th</sup> year	:	4.5 %
5 <sup>th</sup> year	:	5%
5. Banker Detail's : Name : Brainrootz Labs PVT. LTD.  
Account No : 023283900000632  
IFSC Code : YESB0000232  
Bank Name : Yes Bank
6. Country of Origin : United Kingdom ( U. K )
7. PBG : As Per AIIMS terms & conditions

**Note :** We hereby accepts all terms & conditions as applicable of the A.I.I.M.S, New Delhi

Thanking You

Yours Faithfully,

For Brainrootz Labs Pvt. Ltd.  
For BRAINROOTZ LABS PVT. LTD.

Authorized Signatory  
Manish  
Marketing Manger  
Mob No. +91-9999299871

Dr. P.N. DOGRA  
आचार्य एवं अध्यक्ष / Professor & Head  
गुणरोग विज्ञान विभाग / Deptt. of Urology  
अखिल भारतीय आयुर्विज्ञान संस्थान  
All India Institute of Medical Sciences  
अंसरी नगर, नई दिल्ली / Ansari Nagar, New Delhi-28, INDIA



Lamer House Office  
Lamer Lane  
Wheathampstead  
Herts AL4 8RL  
United Kingdom  
t: +44 (0)1582 614 466  
contact@combat-medical.com

Date: 09/02/2017

**TO WHOMEVER IT MAY CONCERN**  
**PROPRIETARY ARTICLES CERTIFICATE**

This is to certify that M/s. Brainrootz Labs Pvt. Ltd. is authorised to market the Combat Medical Ltd products listed below, in the North and East regions of India:


- **COMBAT BRS System**
- **COMBAT BRS Set/Kit, manufactured by Combat Medical Ltd.**

COMBAT BRS System and COMBAT BRS Set/Kit are Proprietary Items of **COMBAT MEDICAL LTD.**

M/s. Brainrootz Labs Pvt. Ltd., New Delhi is the exclusive authorised distributor for marketing **COMBAT BRS System and COMBAT BRS SET** in the North and East regions of India.

Please feel free to contact M/s. Brainrootz Labs Pvt. Ltd., or COMBAT MEDICAL LTD, for any clarification.


Yours Sincerely

For M/s.   
Authorised Signatory


**Name:** Diana Korda Hewitt  
**Designation:** Head of RAQA  
**Address:** Lamer House Office  
Lamer Lane  
Wheathampstead  
AL4 8RL  
England, UK



Company Number 0087673  
VAT Number 155873870  
[www.combat-medical.com](http://www.combat-medical.com)

Page 2 of 2





  
आचार्य प्रोफेसर & डॉ.  
मूत्रपेश विज्ञान विभाग/Deptt. of Urology  
अखिल भारतीय आयुर्विज्ञान संस्थान  
All India Institute of Medical Sciences  
अंसरी नगर, नई दिल्ली/Ansari Nagar, New Delhi-29, INDIA

## **Specification for Combined Antineoplastic Thermotherapy Bladder Recirculation system for Ca Bladder.**

- Unit should deliver Hyperthermic Intra Vesicle Chemotherapy at 43 degree Celsius.
- Unit should use an external dry system and have aluminium heat exchange to ensure efficient heat transfer with temperature control plus/minus 1 degree Celsius.
- Unit should have facility for removal & disposal of Mitomycin (MMC) at the end of treatment from the patient.
- Unit should be portable, robust and user friendly and have Touch screen for easy settings.
- Unit should require minimal set up
- Should have pressure sensor and tube detection
- Unit should have Peristaltic pump for accurate and continuous recirculation and flow
- Unit should maintain temperature of recirculation of fluid ( MMC + Saline ) at 43degree Celsius.
- Should have safety alarms for high and low temperature and overpressure, requiring no continuous monitoring
- Should be CE/US FDA Approved.
- Should quote all the necessary consumables required for recirculation.
- The comprehensive warranty will be 5 years (including all spares and labour) from the date of satisfactory installation of equipment. Also quote rates for comprehensive CMC (including all spares and labour) for 6<sup>th</sup> to 10<sup>th</sup> year, after expiry of warranty period. Cost of spares, accessories and consumables should also be quoted separately.

*[Handwritten signatures and initials]*