



**DEPARTMENT OF MICROBIOLOGY  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES**

S. No: IX/N-1405/2014-15

Dated: 08.08.2014

*A short rate enquiry (LIMITED TENDER) is floated/invited for the purchase of below mentioned item as per the specifications.*

Name of item	Qty.
FILTRATION UNIT	01

**Specifications:**

1. It should be made up of autoclavable Polycarbonate/stainless steel material.
2. It should be of Filter Funnel type with cover.
3. The filtration capacity should be upto 1L.
4. It should attach to 1000ml flask through vacuum pump.
5. It should accommodate disposable membrane filter of 0.45uons and 0.2uons with 47mm dia.
6. The unit should come with minimum of 2 year warranty.

**Terms & Conditions:-**

1. Firm should submit VAT registration certificate.
2. Authority letter from manufacturer/principal should be annexed with quote.
3. The firm is requested to submit the undertaking on the firm's letterhead that the firm has **not been black listed** in the past by any hospital/organization.
4. The firm is requested to submit the undertaking on the firm's letterhead that the firm has **no vigilance case/CBI case pending against him/suppliers.**
5. The firm is requested to submit Fall clause undertaking on the firm's letterhead.
6. Firm should clearly mention the **brand name** for the product.
7. Literature of the original catalogue of the product.
8. The sealed quotation should reach in **Room# 2080**, Microbiology department under two bid system on or before **16/08/2014 upto 12:00 noon** after entry being made in register.
9. The Technical Bids may be opened on **16/08/2014 at 12:30 pm** in the undersigned office.

Tender should be submitted in **TWO BID SYSTEM** containing two parts as detailed below:-

**Part I : Technical Bid in one sealed envelope**

**Part II : Price Bid in one sealed envelope**

The sealed envelopes should then be put in one outer envelope indicating thereon:

- a) Reference No. of the short rate enquiry
- b) Item Name
- c) Name of the Firm

*Those firms who fulfils/submit pre-qualification documents and meet the technical specifications shall be called for demonstration of the equipment, if felt necessary. The final technical selection will be made only after satisfactory demonstration of quoted item/model.*

*(Handwritten signature)*  
**(STORE OFFICER)**

**DR. IMMACULATA XESS**  
Professor  
Department of Microbiology  
A.I.I.M.S., New Delhi-110029