

Bid Corrigendum

GEM/2025/B/6247242-C3

Following terms and conditions supersede all existing "Buyer added Bid Specific Terms and conditions" given in the bid document or any previous corrigendum. Prospective bidders are advised to bid as per following Terms and Conditions:

Buyer Added Bid Specific Additional Terms and Conditions

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. Make in india specific authorisation certificate needs to be enclosed.
4. OPTION CLAUSE: The Purchaser reserves the right to increase or decrease the quantity to be ordered up to 50 percent of bid quantity at the time of placement of contract. The purchaser also reserves the right to increase the ordered quantity up to 50% of the contracted quantity during the currency of the contract at the contracted rates. The delivery period of quantity shall commence from the last date of original delivery order and in cases where option clause is exercised during the extended delivery period the additional time shall commence from the last date of extended delivery period. The additional delivery time shall be $(\text{Increased quantity} \div \text{Original quantity}) \times \text{Original delivery period (in days)}$, subject to minimum of 30 days. If the original delivery period is less than 30 days, the additional time equals the original delivery period. The Purchaser may extend this calculated delivery duration up to the original delivery period while exercising the option clause. Bidders must comply with these terms.
5. Buyer uploaded ATC document [Click here to view the file.](#)
6. Buyer Added text based ATC clauses

CORRIGENDUM

Sub.: Purchase of "Ophthalmic OT Table (General Anaesthesia) - 05 Nos." in Dr. R.P. Center thereof.

**Ref. ATE No. : 05/GeM-Category/RPC/SSK/2025- 26
(GEM/2025/B/6247242, dt. 19/05/2025)**

It is submitted that a tender of "**Ophthalmic OT table (General Anaesthesia) - 05 Nos.**" vide ATE No. : 05/GeM-Category/RPC/SSK/2025- 26 (GEM/2025/B/6247242, dt. 19/05/2025) was published through GeM portal. In view of representations received from prospective bidders, tendered specifications of aforesaid equipment had been amended, as per detail given below:

1	Multipurpose Operating Table suitable for ophthalmic Surgeries in supine position.
----------	------------------------------------------------------------------------------------

2	The table should have solid Stainless Steel frame Construction.
3	<p>The table should have Electro Hydraulic Dual Telescopic column with height of 600 mm-850 mm.</p> <p><u>MAY BE READ AS UNDER</u></p> <p>“ The table should have Electro Hydraulic Dual Telescopic column with height of 700 mm-1000 mm”.</p>
4	The Table should have memory function to achieve at least 4 different positions which can be set individually. Positions can be selected manually by foot paddle and by remote control.
5	The Table should have 4 easy drive castors with electro hydraulic locking control, with connecting axle & lift-brake system.
6	The Table should have 24V DC battery pack operation. The Unit should run for 1-2 weeks when fully charged with battery indicator.
7	The Table should have Hand held remote control with at least 4 memory functions.
8	The Table should be able to achieve lying position and should have head up, lateral tilt, trendelenburg and reverse trendelenburg option with detachable armrests (2 Nos with each table)
9	<p>The Surgical Table should have moulded and antistatic Mattress with no seams and easy simple</p> <p>Fixing. Mattress should be of high quality that spans tabletop break for improved patient support. Mattress must be Latex, & CFC, which makes it 100% hygienic with seamless finish and can be washed down completely.</p>
10	<p>The Surgical Table should be in three sections (head section, back section, combined middle & leg section) for conducting surgery.</p> <p><u>MAY BE READ AS UNDER</u></p> <p>“The surgical table should be in three/four sections (head section, back section, combined middle & leg section) for conducting surgery”.</p>

1 1	Head section should be custom made according to the user requirement with a length of 350 mm and in hexagonal shape instead of rectangle shape to provide enough space to surge on to operate from temporal side. Head section should allow enough unobstructed space so that surgeon's legs can be accommodated beneath the head section without any discomfort.
1 2	Head section should be able to lift up & down with one knob which should be present beneath of the head section & should not obstruct surgeon leg during surgery in sitting position (Tilt is required).
1 3	<p>Total Length : Range 1820 mm to 1920 mm</p> <p>Width without side rail : Approx. 500 mm to 510 mm</p> <p>Width with side rail : approx. 540 mm to 560 mm</p> <p>Patient Weight Capacity : Upto 250 Kg <u>MAY BE READ AS</u> "Patient Weight Capacity : Atleast 250 Kg"</p> <p>Max Load of Head Section: 30-35 kg</p> <p>Max Load of Foot Section : 60 - 70 kg</p> <p>Height adjustment: 600 mm to 885 mm <u>MAY BE READ AS</u> "Height adjustment: 700 mm to 1000 mm".</p> <p>Head section adjustability: +25 deg to -15deg. <u>MAY BE READ AS</u> "Head section adjustability: Atleast +25 deg to -15deg".</p> <p>Back Section adjustability 0 to + 70 deg. <u>MAY BE READ AS</u> "Back Section adjustability Atleast 0 to + 70 deg"</p> <p>Trendelenburg and reverse trendelenburg position: -20/ +20 degree <u>MAY BE READ AS</u> "Trendelenburg and reverse trendelenburg position: Atleast -20/ +20 degree"</p> <p>Lateral tilt ; -18/ +18 degree <u>MAY BE READ AS</u> Lateral tilt ; Atleast -18/ +18 degree</p>
1 4	The equipment should be CE (European)/ US-FDA / Equivalent Indian Standard Certificate like CDSC.

1	Mounting Brackets/ Clamps for hand rest -2 with each table
5	Arm Rest-2 Nos. <u>MAY BE READ AS</u> “Detachable Arm Rest-2 Nos.
	Mounting Brackets/ Clamps for anaesthesia frame - 2 with each table
	Anaesthesia Frame (L shape)
	Gel Pad:
	a) Head Ring (Paediatric) -2 Nos.
	b) Head Ring (Adult) - 2 Nos.

It is informed to all prospective manufacturers/ supplier of aforesaid equipment to read the above amendment only and other conditions in bid remain same.

Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.
3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for [attached categories](#), trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.

14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

*This document shall overwrite all previous versions of Bid Specific Additional Terms and Conditions.

[This Bid is also governed by the General Terms and Conditions](#)