

DR. RAJENDRA PRASAD CENTRE FOR OPHTHALMIC SCIENCES
AIIMS, ANSARI NAGAR, NEW DELHI - 110 029
(STORE SECTION)

DATED: 10.09.2025

CORRIGENDUM

Sub.: Purchase of “ND YAG Laser Machine – 01 No.” in Dr. R.P. Center thereof.

Ref. ATE No. : 19/CPPP-GT/RPC/SSK/2025-26 (2025_AIMSD_873125_1)

It is submitted that a Global Tender of “ND YAG Laser Machine – 01 Nos.” vide No. 19/CPPP-GT/RPC/SSK/2025-26 (2025_AIMSD_873125_1) was published through CPP portal. In view of representations received from prospective bidders, tendered specifications of aforesaid equipment had been amended, as per detail given below:

Tendered Specifications			Amended Specifications
1	Laser Source	Q – Switched ND YAG	Remain Same
2	Wavelength	1,064 nm	Remain Same
3	Pulse Width	3 ns	3 – 5 ns
4	Optical Breakdown in air	2.0 mj or less	Remain Same
5	Pulse repetition rate	3 Hz (Single) / 1.5 Hz (Burst)	Remain Same
6	Output Energy	0.3 to 10.0 mj/ pulse	Remain Same
7	Burst Mode	1,2 and 3 pulse per trigger	Remain Same
8	Spot Size	8 µm	Remain Same
9	Cone Angle	16°	Remain Same
10	Focus Shift	0 to + 500 µm continuously variable	Remain Same
11	Aiming Beam	635 nm/ OFF, 0.5 to 25 µm	Remain Same
12	Motorized Rotable Aiming Beam	Aiming beam rotation of 360 Degree to avoid corneal opacities to achieve more accurate focussing.	Remain Same
Slit Lamp			
13	Illumination	LED Lamp	Remain Same
14	Magnification (Field of View)	5 x (40.7 mm). 8 x (25.7 mm), 12.5 x (16.1 mm), 20 x (10.1 mm), 35 x (6.4 mm)	10 – 25 x with atleast 3 steps
15	Power Supply	220V, 50/60 Hz	Remain Same
16	Dimensions	346 (W) x 422 (D) x 577 (H) mm/ 17 Kg	Deleted
17	Optional Accessories	Foot switch, stand for control for box, safety goggles, illumination tower (tilting)	Illumination Tower

It is informed to all prospective manufacturers/ supplier of aforesaid equipment to read the above amendment only and other conditions in bid remain same.

Stores Officer
(Store Section, Dr. RPC)