



**All-India Institute of Medical Sciences  
Ansari Nagar, New Delhi-29.  
(STORES SECTION (DO))**

Ref. No.XX-91/Stores (DO)/PMSD./2016-17/FSC-IV.

Dated: 14th Spetember, 2016


**Subject:** Purchase of "High Flow Oxygen Devices-02 Nos" for the department of PMSD at AIIMS, New Delhi-29 on proprietary basis- Inviting comments thereon.

\*\*\*\*\*

The Institute is in the process to purchase "High Flow Oxygen Devices-02 Nos" for the department of PMSD at AIIMS, New Delhi-29 from M/s.Fisher & Paykei Healthcare India Pvt. Ltd., Bangalore on proprietary basis. The proposal submitted by M/s. Ved Med Software & Trading Pvt. Ltd., New Delhi and PAC certifications are attached.

The above documents are being uploaded for open information to submit objections, comments, if any, from any manufacturer regarding proprietary nature of the equipment/item within issue of 14 days giving reference No.XX-91/Stores (DO)/PMSD./2016-17/FSC-IV. The comments should be received by office of Stores Officer (DO), Store Section, Animal House Building, Near Biotechnology Building at AIIMS on or before 01-10-2016 upto 12.30 p.m., failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

Yours faithfully,

  
14.9.16  
SR.STORES OFFICER

**Encl:** Related documents enclosed.

ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI-110029

**PROPRIETORY/ SPECIFIC BRAND GOODS CERTIFICATE**

1. Item/Type/Model No. required alongwith specification High Flow Oxygen Device
2. Is the item a spare parte attachment or accessory for an existing equipment. Yes
3. Name of the manufacturers/supplier of the item proposed by the Indentor. M/s Ved Med Software & Trading Pvt. Ltd, B-29, Kailash Colony, New Delhi - 110048,
4. Are the sole manufacturers/sold distributors of the item. Yes (as the certificate provided by the firm)
5. Is there any other item with similar/ equivalent specification available in the market to meet the job requirement envisaged, if the answer is yes, why the same can't be procured. Demanding officer should bring out comparative functional advantages/cost effectiveness of the recommended item from these offered by other. No (High Flow Oxygen Device requirements of the specifications of the item needs for the Department of PMSD)
6. What were the efforts made to locate alternative source of supply or use other substitutes. ---
7. Why open/limited tender can't be resorted to, for locating alternative source. As the firm is sole Proprietor
8. Are the proprietary items certifying that the rates are reasonable or not. Yes the rates are reasonable
9. Any other justification for procuring item from single source. No

Signature of Indentor  
(Demanding Officer)

Additional Professor  
Deptt. of Pulmonary Medicine & Sleep Disorders  
All India Institute of Medical Sciences

COUNTERSIGNED  
(Head of the Department)

I certify that the item at Sr. No. 1 above is required to be procured on single tender basis as the supply is definitely known/ the specified brand proposed was advantages in meeting our functional requirements and limited tender system could be dispensed with as they would serve no useful purpose in this particular case.

(Strike out whichever is not applicable)



# Fisher & Paykel HEALTHCARE

Fisher & Paykel Healthcare Limited  
15 Maurice Paykel Place, East Tamaki  
P O Box 14 348, Panmure  
Auckland, New Zealand  
Telephone: +64 9 574 0100  
Facsimile: +64 9 574 0158  
Website: www.fphcare.com

F&P/PROP/AIRVO/10/13-14/190  
Date: 12 November 2013


## TO WHOM IT MAY CONCERN






### Proprietary Certificate for The AIRVO 2 Humidifier

This is to certify that the AIRVO 2 Humidifier with integrated flow generator is proprietary technology owned and manufactured solely by Fisher & Paykel Healthcare Limited, New Zealand.

The AIRVO 2 Humidifier with integrated flow generator delivers high flow warmed and humidified respiratory gases to spontaneously breathing patients through a variety of patient interfaces.

Yours sincerely,  
For Fisher & Paykel Healthcare Ltd,

  
Authorized signatory.

 x    
 





# Fisher & Paykel HEALTHCARE

Fisher & Paykel Healthcare Limited  
15 Maurice Paykel Place,  
East Tamaki  
P O Box 14 348, Panmure  
Auckland, New Zealand  
Telephone: +64 9 574 0100  
Facsimile: +64 9 574 0158  
Website: www.fphcare.com

19 August 2013

To whom it may concern,

## Proprietary Certificate for the AIRVO™ 2 Humidifier

This is to certify that the AIRVO 2 device, which delivers high-flow humidified respiratory gases with the aid of an integrated high-flow generator and oxygen blending system, is proprietary technology owned and produced by Fisher & Paykel Healthcare Ltd, New Zealand.


Yours faithfully,



Kevin O'Donnell  
Product Group Manager  
Fisher & Paykel Healthcare Ltd  
Auckland, New Zealand



13.08.19 - AIR - India Proprietary Certificate



**DEPARTMENT OF PULMONARY MEDICINE & SLEEP DISORDERS  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES**

**Item: High flow oxygen device**

**Quantity: 2**

**Estimated Cost (5 Lacs each): 10 lacs**

**Specifications:**

1. Gas flow rate: 2-55 lit/min or better
2. Flow setting: with facility of increment of 1 lit/min
3. Humidification: It should have integrated humidifier
4. Interfaces: It should be capable of providing high flow oxygen through – nasal cannula, tracheostomy and mask

**Accessories**

- a. Interfaces: (for each device)
  1. Nasal cannula – 300 (small, medium and large, 100 each)
  2. Tracheostomy interface – 100
  3. Mask interface adapter – 100
- b. All accessories needed for Pole mounting of the system
- c. Disinfection kit – 10
- d. Air filters – 500
- e. Power supply –
  1. Power input to be 220-240V AC, 50 Hz fitted with Indian Plug
  2. Battery back-up of atleast 30min

