

ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI - 110029

PROPRIETARY/SPECIFIC BRAND GOODS CERTIFICATE

1. Items/type/model No. Required along with Maintenance specification **WRIST SUPPORTING SYSTEM FOR OPERATING**
2. Is the item a spare part attachment Or accessory for an existing equipment **NO required for surgeon's comfort**
3. Name of the manufactures/supplier Of the item proposed by the indenter **ORITAB, GERMANY long duration surge**
4. Are they sole manufactures/ Sole distributors of the items **YES**
5. Is there any other item with similar/equivalent Specification available in the market to meet The job requirement envisaged. If the answer is yes Why the same can't be procured. Demanding officer should bring out comparative Functional advantages/cost effectiveness of the Recommended item from these offered by other **NO**
6. What were the efforts made to locate alternative Source of supply or use other substitutes **YES, THROUGH INTERNET AND EXISTING VENDORS**
7. Why open/limited tender can't be resorted to, For locating alternative source. **SINGLE COMPANY we have placed**
8. Are the proprietary certify that the rates are Reasonable or not **YES a tender, only as the bid by same company, her request for proprietary product (Tender no. 100/cn**
9. Any other justification for procuring item for single **NS/14-15/ST)**

*Abhinav*  
**Officer In Charge**  
**Neurosurgery Store**  
Prof. P. Sarat Chandra  
Professor  
Dept. of Neurosurgery  
Neuro sciences Centre  
All India Institute of Medical Sciences, New Delhi-29

*Dr. B. S. Sharma*  
**Countersigned**  
**(HOD, Neurosurgery)**  
**DR. B. S. SHARMA**  
Professor & Head  
Neurosurgery  
All India Institute of Medical Sciences,  
New Delhi-110 029, INDIA

I certify that the items at Sr. No. Above is required to be procured on a single tender basis as the source of supply is definitely know/the specified brand proposed was advantageous, meeting our functional requirements and limited tender system could be dispensed with as they would serve no useful purpose I this particular case.

(Strike out whichever is not applicable)



Date : 06-07-2015

To,  
The Director,  
All India Institute of Medical Sciences,  
Ansari Nagar,  
New Delhi-110029

Sub.: Manufacturer's Authorization Letter & PAC

Ref.: Your TE document No. 100/CNC/NS/14-15/ST FOR WRIST SUPPORT  
DUE ON - 29.01.2015

Dear Sir,


With reference to the above mentioned tender, we are pleased to submit Manufacturer's Authorization Letter and Proprietary Article Certificate.

Kindly acknowledge the receipt of the same.

Thanking You,

With regards,

**FOR INDIAN SURGICAL SYSTEMS**

  
AUTH. SIGNATORY

**ORITAB**

- Mechanische Bearbeitung
- Druckgussteile
- Kunststoffteile
- Blechteile

**SECTION – XIV**  
**MANUFACTURER'S AUTHORISATION FORM**

To  
The 'Director'  
All India Institute of Medical Sciences  
Ansari Nagar,  
New Delhi-29  
India.

Dear Sirs,

Ref. Your TE document No 100/CNC/NS/14-15/ST

We, Oritab AG, Schulstrasse 11, 8542 Wiesendangen, Switzerland who are proven and reputable manufacturers of OP Arm Rest (name and description of the goods offered in the tender) having factories at Schulstrasse 11, 8542 Wiesendangen, Switzerland, hereby authorise Messrs Indian Surgical Systems, B-2/2, IIInd Floor, Double Storey, Ramesh Nagar, New Delhi-110015 to submit a tender, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.

Yours faithfully,

Wiesendangen, January 23, 2015

  
.....  
Bruno Ammann, Director

Wiesendangen, 4. Juli 2015

Dear Sir,

This is to certify that Wrist Support is our proprietary product and no other company can supply the same.

Thanking you and assuring you of our best attention at all times,

Sincerely yours,  
ORITAB AG



Bruno Ammann  
Director

1/2

## WRIST SUPPORTING SYSTEM FOR OPERATING CHAIRS

UNITS REQUIRED: 7, COST PER UNIT: 5 LAKHS, TOTAL COST: 35 LAKHS

1. The manufacturer should be of international repute having current and valid USFDA certification or CE approval for the product
2. The wrist support system should be meant for supporting the surgeons wrist and forearms close to the surgical field during performance of microsurgical procedures so as to prevent fatigue to the surgeon during long duration of surgeries
3. The system should be a complete unit portable and should be light and compact enough to permit rapid shifting and adjustment (on the operating room floor) by the operating surgeon without breach in sterility.
4. The system should not require attachment or fixation to the operating bed or head clamp for operation.
5. The system should be of adjustable height, so that the height could be easily varied intra-operatively by foot operated controls
6. The system should permit tilting of the wrist support system for about 20 degrees in all directions (the extent of mobility should form an inverted cone around the base of the system and the system should be stable against toppling over in this permissible cone).
7. The height should be adjustable between  $90-120 \pm 10$  cm to allow its use for all patient positions.
8. During operation, it may be deactivated and adjusted within an area of 12-140 mm.
9. The arm of the device may be removed and gas sterilized.
10. **The desired height of the arm rest can be obtained by lifting or lowering the supported arms.**
11. The manufacturer can quote a MRI compatible equipment if available.
12. Preferably, the system should not be mechanical and should not require external electric power/ rechargeable

proprietary  
Item. please  
put on website  
20/2/15

2/2

battery for operation.

STORE SECTION  
C.N. CENTRE

Dated: 05.02.2015

Subject: Support system for operating chairs for the Department of Neurosurgery

*20/2/15*

An order No. 100/CNC/NS/14.15/ST was floated in AIIMS website, leading newspaper for the purchase of above mentioned item against the request of Department of Neurosurgery as per their technical specification. In this regard against this tender only single firm M/s Indian Surgical System has submitted their offer. In this regard as per memorandum no. Policy/CPO(AIIMS)/2012 dated 01.10.2013 issued by Chief Procurement Officer regarding Single Tender (received single tender or single short listed bidder).

i) Single Tender (received single tender or single short listed bidder) is required to be re-tendered with the revised broad based specifications. The Store Purchase Committee (Director Office) has also recommended that bringing such procurement proposals where the department has not submitted detailed justification to consider the case in Store Purchase Committee (SPC) not only result in delay but also wastage of manpower & stationary as in any case of SPC straightway recommends re-tendering in such cases.

Accordingly, the user Departments may straightway take action to re-tender and prepare the revised broad based specifications under advice to Store Section to take necessary action for releasing the Earnest Money deposit and closing the bid. In case, any Department considers that no useful purpose will be served by re-tendering and submits detailed justification to consider the case on single shortlisted bid basis, only such cases will be put up by Store Section where the SPC for examination and recommendations.

ii) Further, in respect of procurement cases where the recommendations are made by the SPC for accepting the Single Tender (received single tender or single shortlisted bidder) after the first tender on the basis of detailed justification given by the Department or after re-tender with revised broad based specifications, in such cases before opening the price bids, the Finance Division's concurrence will be obtained before the proposal is put up to the Director (AIIMS) / Chiefs of the Centers for approval in respect of cases falling under their delegated powers.

Submitted for information & necessary action please.

*HL*

(Asstt. Stores Officer)

Prof. Sarat Chandra, Dept. of Neurosurgery

*No objection*

*20/2/15*

*Sarat*

**CARDIO THORACIC & NEURO SCIENCES CENTRE  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES**

Dated: 03.08.2015

Subject: Purchase of Wrist Supporting System from M/s Indian Surgical System an authorized distributor of M/s ORITAB, Germany on proprietary basis for the department of Neurosurgery.

The requirement of above mentioned item received from Department of Neurosurgery. The competent authority recommended to purchase above mentioned item as proprietary item of M/s ORITAB, Germany. The PAC submitted by the department of Neurosurgery & Proprietary certificate provided by the firm are attached & uploaded on website.

The above documents are being uploaded for open information to submit objections, comments, if any, from any manufacturer regarding proprietary nature of the equipment/items within 15 days from the date of issue/uploading of the notification. The comments should be sent to Asstt. Stores Officer, C.N. Centre at AIIMS on or before **21.08.2015 up to 4:30 P.M.**, failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

Yours faithfully

  
(Asstt. Stores Officer, CNC)

**Encl: Related documents enclosed.**

- 1. PAC Certificated enclosed.**
- 2. Proprietary certificate.**