

DEPARTMENT OF ENDOCRINOLOGY & METABOLISM  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
NEW DELHI -110029

Dated: 16/06/2026

**Subject : Purchase of ProFlex PCR System on Proprietary Basis**

This is the approval of purchase of ProFlex PCR system on proprietary basis from M/s Cella Solutions. It is a proprietary product of M/s Thermo Fisher Scientific (Life) and M/s Cella Solutions is authorized distributor for the same. Total Cost of instrument is Rs. 983825.00. The is to be uploaded on AIIMS website for objections or response for 15 days for proprietary purchase.

Thanking you,



Dr. Nikhil Tandon  
Prof. & Head of the Department  
Dept. of Endocrinology & Metabolism

डॉ. निखिल टंडन एम.डी., पी.एच.डी.  
Dr. NIKHIL TANDON M.D., Ph.D.  
आचार्य एवं अध्यक्ष/Professor & Head  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

डॉ. रवींद्र गोस्वामी  
Dr. RAVINDER GOSWAMI  
M.D., D.M. (Endo), FASc., FNAsc. FAMS  
आचार्य/Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029



डा० विवेका पी. ज्योत्सना  
DR. VIVEKA P. JYOTSNA  
एम.डी., डी.एम. (एन्डो)/M.D., D.M. (Endo)  
आचार्य/Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA; Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

## Specification of THERMAL CYCLER WITH 3 x 32 wells Block and 96 well block

- The system should have 3 x 32 block with independent control.
- The system should have option of upgrading with interchangeable 2 x 96 well (0.2 ml) block, 2 X 96 flat block, 2 x 384 well Block.
- The system should support volumes ranging from 10–80 µL/tube.
- Minimum of three different experiments to be run at same time and have the ability to set up PCR with a specific temperature differential of up to 2 temperature zones per block (5°C zone-to-zone) in 3x32 block and has 6 temperature zones in 96-well, 0.2 ml.
- Programmable Heat Lid should be programmable between 50°C to 105°C to perform oil free PCR.
- The system should provide for both Standard and Fast run modes in a single instrument with run times of < 1Hr. 40 Mins on Standard mode and < 35Mins in fast mode.
- The system to support block temperature range starting from 0°C to 100.0°C.
- The system should have a
  - ✦ Temperature Accuracy  $\pm 0.25^{\circ}\text{C}$  (35–99.9°C)
  - ✦ Temperature Uniformity  $< 0.5^{\circ}\text{C}$  (20 sec after reaching 95°C)
  - ✦ Block ramp rate:  $6.0^{\circ}\text{C}/\text{sec}$  (for 3x32 Block and 0.2ml 96 Well Block)
  - ✦ Sample Ramp Rate:  $4.40^{\circ}\text{C}/\text{sec}$
- The system should have state-of-the-art graphical user interface (8.4" color TFT LCD) for easy programming and quickly checking the run status at the bench.
- Simulation mode to make the transition from previous thermal cyclers to the PCR System for simple, accurate, and efficient operation.
- The system should be Wi-Fi enabled and remote connected. The thermal cycler should be cloud-enabled instrument, allowing one secure private access with an account.
- The system to have provision to save up to 800 protocols in the instrument. The system should allow easy product/ software updates via USB port.
- The system to have provision to create separate user profiles protected by passwords.
- The system should have a USB port to transfer methods from one machine to another.
- The system should have 2 years warranty.
- The system should be supplied with an extra 96 well block as an accessory

डॉ. राविवर्द्धन गोस्वामी  
Dr. RAVINDER GOOSWAMI  
M.D., D.M. (Endo), F.A.S.C., F.N.A.S.C. F.A.M.S.  
आचार्य/Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

डॉ. निखिल टंडन एम.डी., पी.एच.डी.  
Dr. NIKHIL TANDON M.D., Ph.D.  
आचार्य एवं अध्यक्ष/Professor & Head  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

डॉ. आशु  
Dr. ASHU  
आचार्य/Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-29


डॉ० विवेका पी. ज्योत्सना  
DR. VIVEKA P. JYOTSNA  
एम.डी., डी.एम. (एन्डो)/M.D., D.M. (Endo)  
आचार्य/Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR : NEW DELHI: 110029

DATED- 16-06-26

PROPRIETRY/SPECIFICATION BRAND GOODS CERTIFICATE

1. Item/Type/Model No. required along with specification. Pro-Flex PCR System, 3X 32 well
2. If the item a spare part attachment or accessory for an existing equipment. No
3. Name of the manufacturer/supplier of the item proposed by the indenter. M/s Thermo Fisher Scientific (LIFE) is the sole manufacturer and M/s Cella Solutions is authorized distributor
4. Are they sole manufacturer/sold distributor of the item Yes - sole manufacturer
5. If there any other item with similar /equivalent specification available in the market to meet the job requirement envisaged. If the answer is yes, why the same can't be procured. Demanding officer should bring out comparative functional advantages /cost effectiveness of the recommended item from these offered by other. No
6. What were the efforts made to located alternative source of supply or use other substitutes. M/s Thermo Fisher Scientific (LIFE) is the sole manufacturer
7. Why open/limited tender can't be resorted to, for locating alternative source. Because no other firm manufacture this instrument.
8. Are the proprietary items certifying that the rates are reasonable or not. Yes proprietary items & the rates are reasonable.
9. Any other justification for procuring item from the single source. This is the only instruments with required specifications.

  
Signature of Indenter  
(Demanding Office)

I certify that the item at Sr. No. 1 above is required to be procured on single tender basis as the source of supply is definitely known / the specified brand proposed was advantageous in meeting our functional requirement and limited tender system could be dispensed with as they would serve no useful purpose in this particular case.

(strike out whichever is not applicable)

डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA; Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29 - 2 -

डा० विवेका पी. ज्योत्सना  
DR. VIVEKA P. JYOTSNA  
एम.जे., डी.एम. (एन्डो)/M.D., D.M. (Endo)  
आचार्य/Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

(Countersigned)  
(Head of the Department)  
डॉ. निखिल टंडन एम.डी., पीएच.डी.  
Dr. NIKHIL TANDON M.D., Ph.D.  
आचार्य/Professor & Head  
अन्तः स्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

DEPARTMENT OF ENDOCRINOLOGY & METABOLISM  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES (AIIMS)  
NEW DELHI -110029

To

The Accounts Officer,  
Budget Section,  
A. I. I. M. S.,  
New Delhi - 110029.

Dated :- ...16/06/26..

Sir,

Please book the funds to the tune of Rs. ..983825.00/-...

(Rupees ....Nine Lakhs Eighty Three Thousand Eight Hundred and Twenty Five rupees only.....)

against the purchase order No. ...034/E&M/26-27....

From ..M/s CELLA Solutions....

For the Department of Endocrinology and Metabolism under the Head **Machine & Equipment Plan -2026 - 2027** , for the purchase of ...**PeoFlex PCR System, 3X 32 well**.

Certified that we members of the purchase committee are jointly and individually satisfied that goods recommended for purchase are of the requisite specification and quality, priced at the prevailing market rate and the supplier recommended is reliable and competent to supply the goods in question.

It is certified that items mentioned in supply order are not found in search on GEM to our best knowledge.

Thanking you,

Head of the Department  
Dept. of Endocrinology & Metabolism

डॉ. निखिल टंडन एम.डी., पी.एच.डी.  
Dr. NIKHIL TANDON M.D., Ph.D.  
आचार्य एवं अध्यक्ष/Professor & Head  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA, Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

डा० विवेका पी. ज्योत्सना  
DR. VIVEKA P. JYOTSNA  
एम.डी., डी.एम. (एन्डो)/M.D., D.M. (Endo)  
आचार्य/Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

**SUPPLY ORDER**  
**ALL INDIA INSTITUTE OF MEDICAL SCIENCES (AIIMS)**

Registered-AD

S. O. No. **034/E&M/26-27**

Ansari Nagar East,, New Delhi -110029

M/s CELLA Solutions  
G3/23, 1st floor , Model Town III, N. Delhi-09

Dated the -  
Date of Delivery -  
or Earlier

Account Code **01-00-102-01-006-000**


Dear Sir,


With reference to your Quotation No. \_\_\_ dated \_\_\_ this office Rate Enquiry/ Tender \_\_\_ dated \_\_\_ . You are requested to arrange supplies of under mentioned items at the price marked against each and deliver the same to the store officer, All India Institute of Medical Sciences, Ansari Nagar, on the terms and conditions stated overleaf.


Sr. No.	Nomenclature	No. of Qty. in fig. & words	Per in full	Rate		Amount		Discount		GST		Total Amt.	
				Rs.	P.	Rs.	P.	Rs.	P.	Rs.	P.	Rs.	P.
1	ProFlex PCR System, 3X32 well	1.0(One)-Nos.	Each	833750.00		833750.00		0.00		@18.00 -	150,075.0	983825.00	
<b>Total</b>											<b>Rs.</b>	<b>983825.00</b>	

**Rs. Nine Lakhs Eighty Three Thousand Eight Hundred and Twenty Five rupees only**

- C. C. to : 1. Store Accounts Section  
2. Central Store Depot.  
3. Firm copy

  
डॉ. निखिल टंडन एम.डी., पी.एच.डी.  
Dr. NIKHIL TANDON M.D., Ph.D.  
आचार्य एवं अध्यक्ष/Professor & Head  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

  
डा० विवेका पी. ज्योत्सना  
DR. VIVEKA P. JYOTSNA  
एम.डी., डी.एच. (एन्डो)/M.D., D.M. (Endo)  
आचार्य/Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-110029  
A.I.I.M.S., New Delhi-110029

  
डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA, Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तःस्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29



**Report ID:** GEM/GARPTS/09062026/57RDRTEG99LC

**Report Name:** Power Flex thermal cyclers, 3x32 well

**Generated By:** Tomer Manjeet Kumar , Department of Health and Family Welfare , Ministry of Health and Family Welfare

**Generated On:** 09/06/2026

**Valid till:** 09/07/2026

### **GeM Availability Report and Past Transaction Summary**

GeM Availability Report and past transaction summary report is generated based on the specifications searched by the Buyer. The specification may be modified appropriately for searching relevant categories on GeM. Buyer may navigate to GeM category page by clicking on the category link to view category specifications and products/services available in the category.

*Order Count and Order Value displayed is on a cumulative basis since GeM inception.*


#### **1. Search String: Power Flex thermal cyclers, 3x32 well**

Search type: Product

Search Result: Category not available on GeM for the text string searched by the buyer.

This is a one-time requirement hence new category creation is not proposed / or requirement is recurring but request for new category creation will be submitted separately post generation of GeMARPTS.

Estimated Value Of Procurement: ₹ 984999.0

  
डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA; Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29




# CELLASOLUTIONS

THE LIFE SCIENCE RESEARCH COMPANION


CELLA SOLUTIONS  
GST NO. 07AEDPV4185L1ZW  
STATE: DELHI (07)  
PAN NO. AEDPV4185L

To, Head of Department Department of Endocrinology All India Institute of Medical Sciences New Delhi-110029	Quotation No.26-27/AIIMS/00532/CS Date: 12/06/26
---	---

S.N.	Cat No.	Description	Size	Offer Price	Qty	Total Price (INR)	GST %	Total Price (INR)
1	4484073	ProFlex™ PCR System, 3 x 32-well  includes 96-well block accessory  (2 years warranty)	each	8,33,750.00	1	8,33,750.00	18%	9,83,825.00
Total (INR)								9,83,825.00
Round Off (INR)								9,83,825.00

<b>Terms &amp; Condition</b> <b>Validity:</b> This quote is valid for 1 month <b>Payment :</b> Against Delivery <b>Delivery:</b> 6-8 weeks after receiving formal order copy	<b>Auth Signatory</b> 
---	--

<b>Special Term &amp; Condition :</b> <b>Shipping Conditions:</b> FOR, Road permit will be provided by buyer's ,if require. The terms and conditions indicated herein including delivery period, prices, products, catalog nos etc may change depending on final confirmation of the same by our principal(s) or on review by us. This Quotation is an indicative form issued in due course of business and not a legal contractual obligation.
---

  
Dr. ASHUTOSH KUMAR ARYA, Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राविकी एवं व्रयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

To,  
Head of Department  
Department of Endocrinology  
AIIMS, New Delhi-110029

12/06/2026

Subject: Fall Clause certificate

Dear Sir,


We Cella Solutions hereby undertake that we have not offered to supply / supplied / are not supplying same or similar products / systems or sub systems at a price lower than that offered against the quotation no. CS/00532/26-27 dated 12/6/26 with similar terms and conditions in respect of any government institute/university during the current financial year and if it is found at any stage that same or similar product/systems or sub systems was supplied by the bidder to any government institute/university at a lower price during the currency of the contract, then that very price will be applicable to the present case and the difference in the cost would be refunded by the bidder to buyer, if the contract has already been concluded.

Regards

**CELLA SOLUTIONS,**

405, 4th Floor, Plot C-1,2,3, P.P. Towers, Netaji Subhash Place,  
Pitampura, Delhi 110034



  
डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA, Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राणिकी एवं तयापचय विभाग  
Dept. of Endocrinology & Metabolism  
अ. भा.आ.सं. नं. 29/A.I.I.M.S., New Delhi-29

To,

12/6/26

Dr Head of Department

Department of Endocrinology

AIIMS, New Delhi-110029

Subject: Non Conviction/Non Blacklist Certificate

Dear Sir/Maam,

In reference to our quotation no. CS/00532/26-27 dated 12-06-26, I Ravi Verma, being the proprietor of M/s Cella Solutions hereby declare that No criminal conviction or pending criminal charges against us.

Also, we hereby declare that we Cella Solutions not been blacklisted/ debarred /banned for the items quoted by us, in the past five years as on the last date of submission of this tender for any reason by any. Central/State Government or any of their organization.

Regards

**CELLA SOLUTIONS,**

405, 4th Floor, Plot C-1,2,3, P.P. Towers, Netaji Subhash Place,  
Pitampura, Delhi 110034



*OK*  
डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA, Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. मा.आ.सं. नई दिल्ली-29/A.I.I.M.S., New Delhi-29

May 16<sup>th</sup>, 2014

**SOLE SOURCE LETTER**

Dear valued Life Technologies Corporation customer,

Thank you for your interest in our Thermal Cyclers products, part numbers listed below.

Part Number	Description
4484071	ProFlex™ 2x96 well Block Module
4484074	ProFlex™ 2xFlat Block Module
4483636	ProFlex™ Base Module
4483637	ProFlex™ 96 well Veriflex Block Module
4484072	ProFlex™ 2x384 Block Module
4483638	ProFlex™ 3x32 Block Module
4484075	ProFlex™ 96 well Veriflex PCR System
4484076	ProFlex™ 2x96 well System
4484078	ProFlex™ 2xFlat Block System
4484077	ProFlex™ 2x384 System
4484073	ProFlex™ 3x32 System

The above products are proprietary technology of Life Technologies, part of Thermo Fisher Scientific (LIFE). LIFE does offer certain third parties limited rights to resell some thermal cycler products to the educational and diagnostic markets.


The Applied Biosystems® ProFlex™ PCR System is covered by at least the IP listed in Exhibit A (Intellectual Property Summary) enclosed herein.

Should you have any questions, please let us know.

Sincerely,



Cain Murphy  
Senior Product Manager  
Endpoint PCR Instruments



डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA; Ph.D.  
सहायक आचार्य / Assistant Professor  
अन्तः स्त्रा की एवं चयापचय विभाग  
Dept. of Endocrinology & Metabolism  
अ. भा.आ.स. 1-29/A.I.I.M.S., New Delhi-29


**Exhibit A: Intellectual Property Summary**

**US Intellectual Property Protection**

1. Title: MULTI-BLOCK PCR THERMAL CYCLER DEVICE  
U.S. Design Patent Application No: 29/438,754  
Status: Pending
2. Title: THERMAL CYCLER DISPLAY SCREEN  
U.S. Design Patent Application No: 29/445,006  
Status: Pending
3. Title: MOBILE APPLICATION USER INTERFACE  
U.S. Design Patent Application No: 29/451,376  
Status: Pending
4. Title: MOBILE APPLICATION USER INTERFACE  
U.S. Design Patent Application No: 29/457,089  
Status: Pending
5. Title: APPARATUSES, SYSTEMS AND METHODS FOR PROVIDING THERMOCYCLER THERMAL UNIFORMITY  
U.S. Provisional Utility Patent Application No: 61/878,464  
Status: Pending


**International Intellectual Property Protection**

1. Title: THERMAL CYCLER DISPLAY SCREEN  
Indian Design Registration Application No: 255624  
Status: Pending
2. Title: THERMAL CYCLER DISPLAY SCREEN  
Indian Design Registration Application No: 255625  
Status: Pending
3. Title: THERMAL CYCLER DISPLAY SCREEN  
Indian Design Registration Application No: 255626  
Status: Pending
4. Title: THERMAL CYCLER DISPLAY SCREEN  
Indian Design Registration Application No: 255627


  
डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA; Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

Status: Pending

5. Title: THERMAL CYCLER DISPLAY SCREEN  
Indian Design Registration Application No: 255628  
Status: Pending
6. Title: THERMAL CYCLER DISPLAY SCREEN  
Indian Design Registration Application No: 255629  
Status: Pending
7. Title: THERMAL CYCLER DISPLAY SCREEN  
Indian Design Registration Application No: 255630  
Status: Pending
8. Title: THERMAL CYCLER DISPLAY SCREEN  
Indian Design Registration Application No: 255631  
Status: Pending
9. Title: THERMAL CYCLER DISPLAY SCREEN  
EU Community Design Registration: 002286104 [0001]  
Status: Issued
10. Title: THERMAL CYCLER DISPLAY SCREEN  
EU Community Design Registration: 002286104 [0002]  
Status: Issued
11. Title: THERMAL CYCLER DISPLAY SCREEN  
EU Community Design Registration: 002286104 [0003]  
Status: Issued
12. Title: THERMAL CYCLER DISPLAY SCREEN  
EU Community Design Registration: 002286104 [0004]  
Status: Issued
13. Title: THERMAL CYCLER DISPLAY SCREEN  
EU Community Design Registration: 002286104 [0005]  
Status: Issued
14. Title: THERMAL CYCLER DISPLAY SCREEN  
Japanese Design Registration Application No: 2013-17830  
Status: Pending

  
डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA; Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29


15. Title: THERMAL CYCLER DISPLAY SCREEN  
Taiwan Design Registration Application No: 102305178  
Status: Pending
16. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257134  
Status: Pending
17. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257143  
Status: Pending
18. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257135  
Status: Pending
19. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257136  
Status: Pending
20. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257137  
Status: Pending
21. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257138  
Status: Pending
22. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257139  
Status: Pending
23. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257140  
Status: Pending
24. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257141  
Status: Pending
25. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 257142

  
डॉ. आशुतोष कुमार आर्या, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA: Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राविकी एवं चयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29




Status: Pending

26. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 002314013 [0001]  
Status: Issued
27. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 002314013 [0002]  
Status: Issued
28. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 002314013 [0003]  
Status: Issued
29. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 002314013 [0004]  
Status: Issued
30. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 002314013 [0005]  
Status: Issued
31. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 002314013 [0006]  
Status: Issued
32. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 002314013 [0007]  
Status: Issued
33. Title: MOBILE APPLICATION USER INTERFACE  
Taiwan Design Registration Application No: 102306215  
Status: Pending
34. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 258512  
Status: Pending
35. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 258513  
Status: Pending

  
डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA, Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्तः स्त्राविकी एवं उपापचय विभाग  
Dept. of Endocrinology & Metabolism  
अ. भा.आ.त्त., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

**ThermoFisher**  
S C I E N T I F I C

36. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 258514  
Status: Pending
37. Title: MOBILE APPLICATION USER INTERFACE  
Indian Design Registration Application No: 258515  
Status: Pending
38. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 02352419 [0001]  
Status: Issued
39. Title: MOBILE APPLICATION USER INTERFACE  
EU Community Design Registration: 02352419 [0002]  
Status: Issued

  
डॉ. आशुतोष कुमार आर्य, पीएच.डी.  
Dr. ASHUTOSH KUMAR ARYA; Ph.D.  
सहायक आचार्य/Assistant Professor  
अन्त : स्त्राविकी एवं त्रयापचय विभाग  
Deptt. of Endocrinology & Metabolism  
अ. भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

**ThermoFisher**  
**SCIENTIFIC**

The world leader in serving science

Invitrogen Bioservices India Pvt.Ltd  
2nd Floor, First Technology Place,  
3 EPIP, Whitefield, Bangalore 560066.  
T+91 80 41785401, 41785400  
F+ 91 080 67745507  
www.thermofisher.com  
CIN: U73100KA2004PTC035330

Ref:15062026-IVGN-BID-AIIMS-AUTHLET-2026  
Dated: 15<sup>th</sup> June 2026

To,  
Head of Department  
Department of Endocrinology  
ALL INDIA INSTITUTE OF MEDICAL SCIENCES (AIIMS)  
Ansar Nagar, New Delhi – 110029

**Ref: Quotation No.26-27/AIIMS/00532/CS dated 12<sup>th</sup> June 2026 for supply of Proflex SKU#4484073**

Dear Sir/Madam,

We, **Invitrogen BioServices India Pvt. Ltd**, having its registered office at Second Floor, First Technology Place, 3EPIP, Whitefield Bengaluru, Karnataka 560066 are the Importer and Supplier of items mentioned as above under Applied Biosystems / Invitrogen brand names hereby authorize **M/s Cella Solutions** having their registered office at G-3/23, 1st Floor, Model Town 3, Delhi-110009 India to participate against the above said enquiry, submit quotes, collect the orders, and supply of items mentioned as above in respect of reference stated above under Applied Biosystems /Invitrogen / Thermo Scientific brand names to you during the fiscal year 2026-2027.

We further authorize **CELLA SOLUTIONS** hereinafter referred to as the **Authorized Distributor**, to directly accept orders for supply, make supply, invoice for such supplies made and collect the payment thereof. The Authorized Distributor is liable to comply with all applicable laws including foreign laws, regulations and policies of Thermo Fisher Scientific Inc. It shall indemnify Thermo Fisher Scientific Inc and its affiliates in case of any claims, losses, penalties, litigation fees, damages caused by breach of the terms of this letter or Distribution Agreement. The authorized distributor shall apply for any preference under the MSME Clause as per his status and certification.

Authorized Distributor acknowledges that the provision by Company of products (including components and spare parts), software, services (including warranty services), technology or intellectual property, including technical information supplied by Company or contained in documents (collectively Items), is subject to applicable export controls of the U.S. government and other jurisdictions, including but not limited to the European Union. These controls may require Company or Authorized Distributor to first obtain a license (or similar requirement) from the relevant authority, or regulatory body. Company shall not be liable to Authorized Distributor for any delay or failure to obtain necessary licenses or approvals. Authorized Distributor shall comply with all applicable export laws and regulations. Authorized Distributor shall not, without first obtaining from the relevant authority or regulatory body any license required to do so lawfully, export or re-export any Item (either directly or indirectly), to: (i) any restricted or embargoed country or any person or organization whose privilege to participate in exports has been denied or restricted by the applicable authority; or (ii) any person or organization who is involved in improper development or use of nuclear weapons, or of chemical/biological weapons (CBW) or missiles, or in terrorist activities. Authorized Distributor agrees not to use any supplied Item in restricted or prohibited activities such as nuclear explosives, unsafeguarded nuclear activities, chemical or biological weapons development, restricted rocket systems, or military purposes. Authorized Distributor will, on request (i) promptly provide written information correctly identifying the end user and end use of any Items (including any information as it may relate to a subsequent transfer of such Items by Authorized Distributor); and (ii) cooperate fully with Company in any official or unofficial audit or inspection arising in respect of the Items under applicable export or import control laws or regulations. Authorized Distributor will ensure that the customers and end users to whom Authorized Distributor re-sells or transfers the Items agree in writing to the provisions of this Section and Authorized Distributor covenants to use its best efforts to enforce such provisions against customers and end users. Authorized Distributor shall indemnify and hold Company harmless from, or in connection with, any violation of this Section by Authorized Distributor or its employees, consultants, agents and/or representatives. In addition, failure of Authorized Distributor to comply with this Section shall be a material breach of this Agreement and shall entitle Company to immediately terminate this Agreement; Company shall be entitled to terminate this Agreement without prior notice if such termination is necessary in order to comply with applicable export laws and regulations. The Authorized Distributor has no right over the Intellectual Property of Invitrogen BioServices India Pvt. Ltd, as well as its affiliate companies and its parent company and the Intellectual Property Rights in the products are only given via a limited license for the purpose of fair and free use of these products for their lifetime. The Authorized Distributor shall not assign any right over the Intellectual Property belonging to Invitrogen BioServices India Pvt. Ltd, as well as its affiliate companies and its parent company to any third party.

**The products mentioned above are 100% imported in India and we are a non-local supplier of the same.**

This authorization is valid for the purpose of the above referred tender only.

Yours faithfully,  
For Invitrogen BioServices India Pvt. Ltd.,

*Vikrant Vishnoi*

Electronically signed by: Vikrant Vishnoi  
Reason: Approver of the Non-GxP document  
Date: Jun 16, 2026 11:19:51 GMT+5.5

Authorized Signatory

Branch Office  
**Bangalore**  
Invitrogen BioServices India Private Limited  
C/O DHL Supply Chain India Pvt Ltd  
NDR Bhadra Estates LLP, Survey No 75-79  
83-85, 101/2-6, 104/2, Makur Rd, Hoskote Tk  
Karibeerahesahalli, Jadigenahalli  
Bengaluru Rural - 562114  
Toll Free Phone No: 1600 419 543

Branch Office  
**Gurgaon**  
Invitrogen BioServices India Private Limited  
372, Udyog Vihar Phase - II  
Gurgaon - 122 016  
Haryana  
T+91 124 6722800  
F+91 124 6722850, 6722851

Branch Office  
**Mumbai**  
Invitrogen BioServices India Private Limited  
403/404 Delphi B Wing,  
Hiranandani Business Park  
Powai, Mumbai-400076  
T+91 22 66803001  
F+91 22 67162244

Branch Office  
**Kolkata**  
Invitrogen BioServices India Private Limited  
Unit No 801, Merin Acropolis  
185/8/1 Rajjanqa Main Road  
Kolkata - 700 107,  
West Bengal  
T+61 33 66887300, Ext-3911

# ProFlex PCR System

The Applied Biosystems™ ProFlex™ PCR System is the only thermal cycler that allows simultaneous remote access by three users to control three independent blocks.



- **Accessible by multiple users**—run three experiments with three independently controlled blocks
- **Flexible block configuration**—accepts five types of thermal blocks for thermal optimization and throughput options
- **Convenient remote access**—connect via the cloud or a mobile device and monitor your instrument from anywhere with a free mobile app

Block format	3 x 32-well, 0.2 mL; independent control	96-well, 0.2 mL	2 x 96-well, 0.2 mL	2 x flat block for chips and arrays	2 x 384-well, 0.02 mL
Features	Run three experiments at once or at different times	Perform complete optimization work with full 96-well Applied Biosystems™ VeriFlex™ Block	High throughput in 96-well format	Highest throughput capability: 8 x 3,072 Applied Biosystems™ OpenArray™ Plate* or 24 x 20K Chip**	High throughput in 384-well format
Max. block ramp rate	6.0°C/sec	6.0°C/sec	3.0°C/sec	1.6°C/sec	3.0°C/sec
Max. sample ramp rate	4.4°C/sec	4.4°C/sec	1.6°C/sec	NA	1.6°C/sec
Temperature accuracy	±0.25°C (35–99.9°C)				
Temperature range	0–100.0°C				
Temperature uniformity†	<0.5°C (30 sec after reaching 95°C)				
Dimensions (H x W x D)	27.2 x 33.0 x 56.5 cm (10.6 x 13 x 22 in.)				
Weight	18.8 kg (41 lb)	18.8 kg (41 lb)	19.3 kg (42.5 lb)	19.6 kg (43.2 lb)	20.4 kg (45 lb)
PCR volume range	10–80 µL	10–80 µL	10–100 µL	33 nL	5–20 µL
Instrument memory	USB, onboard				
Display interface	8.4 in. color thin-film transistor (TFT) LCD				
Power	100–240 V, 50–60 Hz, max. 950 VA				
VeriFlex Blocks	2 temperature zones per block (5°C zone-to-zone)	6 temperature zones, 25°C range (5°C zone-to-zone)	NA	NA	NA
Cat. No.: Complete system (base and block)	4484073	4484075	4484076	4484078	4484077
Cat. No.: Block (includes block only)	4483638	4483637	4484071	4484074	4484072

\* OpenArray Plate is compatible with the Applied Biosystems™ QuantStudio™ 12K Flex Real-Time PCR System

\*\* 20K Chip is compatible with the Applied Biosystems™ QuantStudio™ 3D Digital PCR System

† For 2 x 96-well and 2 x flat block format, temperature uniformity is for each block. Over all temperature uniformity across two blocks is <0.75°C (30 sec after reaching 95°C)

Find out more at [thermofisher.com/proflex](http://thermofisher.com/proflex)

**ThermoFisher**  
SCIENTIFIC