

SPECIFICATIONS FOR GALACTOMANNAN ANTIGEN DETECTION KIT

1. The kit should be highly sensitive to detect galactomannan polysaccharide antigen.
2. The kit should detect galactomannan antigen in human serum and BAL samples.
3. The kit should be in form of individual microwell strip plate with 8 wells per strip.
4. The assay must be monoclonal antibody based and should not cross react to other fungal polysaccharides.
5. The maximum time for an assay should not be more than 3hrs.
6. The assay should be readable in standard ELISA readers with appropriate software.
7. The kit should be FDA/CE approved.
8. The expiry of the kit must be atleast one year.



DR. IMMACULATA XESS
Associate Professor
Department of Microbiology
A.I.I.M.S., New Delhi-110029

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI - 110 029

PROPRIETORY/SPECIFIC BRAND GOODS CERTIFICATE

1. Item/type/Model No. required along with specifications. *Platelia Aspergillus EIA*
2. Is the item a spare part attachment or accessory for an existing equipment? *N.A.*
3. Name of Manufacturers supplier of the item proposed by the indenter. *M/s Bio Rad.*
4. Are they sole manufacturers/sold distributors of the item? *yes*
5. Is there any other item with similar/ equivalent specification available in the market to meet the job requirement envisaged. If the answer is yes, why the same can't be procured. Demanding Officer should bring out comparative functional advantages/cost effectiveness of the recommended item from those offered by other. *- No -*
6. What were the efforts made to locate alternative source of supply or use other substitutes. *- No -*
7. Why open/limited tender can't be resorted to, for locating alternative source. *Proprietary item*
8. Are the proprietary items certifying that the rates are reasonable or not. *reasonable*
9. Any other justifications for procuring item from single source. *Proprietary*

[Signature]
Signature of indenter
(Demanding Officer)

[Signature]
COUNTERSIGNED
(Head of the Department)

**H.O.D**

Deptt. Of Microbiology
A.I.I.M.S. Ansari Nagar
New Delhi-110029

Dated : 27th July 2018

Ref.No. : DMPL/18-19/325

Subject : Quotation

<u>S.No.</u>	<u>Product and Pack</u>	<u>Make</u>	<u>Price(Rs.)</u>	<u>CGST%</u>	<u>SGST%</u>	<u>Total GST Amount</u>	<u>Total Amount with GST</u>
1.	Platelia Aspergillus EIA 96 Test Code No. 62794	Biorad	29970.00	2.5%	2.5%	1498.50	31468.50

Terms and Conditions:

- a. F.O.R. : Yours Hospital Lab.
- b. Supply Period : 2-4 Days
- c. Validity : 90 Days.
- d. HSN Code : 3822

For **Diagaid Mediwares Pvt. Ltd.**



S.P.Singh

(Mg. Director)

Diagaid Mediwares Pvt. Ltd.

Sales Office : M-8, Dewan House, Ajay Enclave, New Delhi-110018, INDIA

Tel. : 91-11-2513 9761, 2513 9762, 91-11-2540 9762, 3259 9762

Fax : 91-11-2513 9762, Cell : 00-98100 62739

Email : diagaid@gmail.com / diagaid@rediffmail.com

CIN NO. : U74899DL1987PTC030061



Bio-Rad
Laboratories (India) Pvt Ltd
(A wholly owned subsidiary)

'Bio-Rad House', Plot # 95-97
Udyog Vihar Phase - IV
Gurgaon-122015, Haryana
Phone: (91)-124-4029300
Fax: 91-124 - 2398115
E-mail: sales.india@bio-rad.com
Tech Support : 1800 180 1224, 09873177477
www.bio-rad.com



Registered Office: Plot No. 1270 Basement, Lai Dora, Village Kapashera,
Opposite Fuz- Food Village, New Delhi - 110037 Phone: 91-11-25065913
CIN No. - U32109DL1996PTC078494

To,
The Director
A.I.I.M.S, Ansari Nagar
New Delhi

Sub. : Authority Letter

Dear Sir,

We wish to inform you that

Diagaid Mediwares Pvt.Ltd.
M-8, Dewan House, Ajay Enclave
New Delhi -110018
Tel. 25139761, 25139762, 25409762

Are our authorized Distributors to
quote our brand of Kits and Controls in yours Hospitals. In the case of an order they are
also authorized to effect supplies on their own invoice and collect payments thereof.

This Authorization is valid for the period upto 31st March 2019

Thanking You,

For Bio-Rad Laboratories(I) Pvt. Ltd.



Authorised Signatory

• Diagnostic Kits • Instruments



• Chemicals • Disposables

H.O.D

Department of Microbiology

A.I.I.M.S. Ansari Nagar

New Delhi

Subject : Our Quotation No. DMPL/18-19/325

This is to certify that Items Quoted are **Proprietary Brand** & Rates charged are our current **minimum rates** and are not more than those charged to Govt. Hospitals, Institutes and departments.

For Diagaid Mediwares Pvt. Ltd.



S.P.Singh

(Mg. Director)

Diagaid Mediwares Pvt. Ltd.

Sales Office : M-8, Dewan House, Ajay Enclave, New Delhi-110018, INDIA

Tel. : 91-11-2513 9761, 2513 9762, 91-11-2540 9762, 3259 9762

Fax : 91-11-2513 9762, Cell : 00-98100 62739

Email : diagaid@gmail.com / diagaid@rediffmail.com

CIN NO. : U74899DL1987PTC030061

BIO-RAD

**Bio-Rad
Laboratories (India) Pvt Ltd**
(A wholly owned subsidiary)

"Bio-Rad House", Plot # 86-87
Udyog Vihar Phase - IV
Gurgaon-122015, Haryana
Phone: 91-124-4029300
Fax: (91)-124-2398115
E-mail: sales.india@bio-rad.com
Tech Support: 1800 180 1224, 09873177477
Website: www.biograd.co.in



Registered Office: plot No. 1270 Basement, Lal Dora, Village Kapashera
Opposite Fun-Food Village, New Delhi-110037 Phone: 91-11-25065913
CIN No :- U32109DL1996PTC078494

To Whom It May Concern

Proprietary Certificate

This is to certify that **Bio-Rad Laboratories Platelia Aspergillus Ag** (Cat # 62794) is intended for use as a microbiology product to monitor the lab testing as per package insert ; are proprietary to Bio-Rad Laboratories Inc. USA, and these products are only manufactured and supplied by Bio-Rad.

Bio-Rad Laboratories India Private Ltd. is a 100% fully owned subsidiary of 4000 Alfred Nobel Drive Hercules, California USA 94547.

For
Bio-Rad Laboratories India Pvt. Ltd.



Sumit

Authorized Signatory