

DEPARTMENT OF MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Prof in charge
Computer facility
AIIMS

27th Feb, 2018


Rate Enquiry No. 01 /Med/DSPC/2017-18

SEALED QUOTATIONS are invited by post for supply of the items as per list attached, which are required for the department of Medicine of the Institute.

The make of the articles offered should invariably be quoted percentage of CGST & SGST or Tin No. May be also being indicated.

The quotations should reach office of Dr. Rita Sood, Prof. & Head, Department of Medicine, conference room no. 3094 A, 3rd floor, Teaching block, AIIMS, New Delhi-110029 before the date as indicated in the enclosed list in a sealed envelope, with written Rate Enquiry Number by courier or post. All quotations should be clearly written or typed. All over writings and entries will quotations disqualified.

The quotations will be opened at 2.30 p.m. on 15th March, 2018 as indicated in the enclosed Performa in the office of the Prof & Head Department of Medicine. AIIMS, New Delhi- 110029 in the presence of the members of committee.


Prof. Naval K. Vikram
Prof-in-charge (Store)
Departmental store committee

DEPARTMENT OF MEDICINE
ALL INDIA INSTITUTE OF MEDICAL SCIENCES

Rate Enquiry No. **01 /Med/DSPC/2017-18**

Dated: 27th Feb, 2018

Date of submission of quotations: 14.03.2018

up to 5.00 p.m.

Date of opening of quotation 15.03.2018

2.30 p.m.

Please find enclosed the following M&E required in the department of Medicine.

S.No.	Nomenclature	Qty Required	Rate Enquiry no.
1.	Hydraulic hand Dynamometer	one	RE no.01/Med/DSPC/2017-18


Prof. Naval K. Vikram
Prof-in-charge (Store)
Department of Medicine


Specification for Hydraulic Hand Dynamometer


Quantity required: 01

Approximate cost: Rs. 75,000/-


1. The Hydraulic Hand Dynamometer should have Dual-Scale readout of grip force in pounds and kilograms.
2. Maximum reading- 200 pounds or 90 kilograms.
3. Peak-hold needle for convenience and easy recording.
4. Accurate and reproducible recording.
5. Comfortable and effective hand grasp.
6. Adjustable handle to accommodate various size hands preferable from 1-3/4 to 3-3/8 inches, in half-inch increments.
7. Firm should quote with warranty for 2 years.



Dr. Pombay
Jorwal



Dr. Manish Sonjeja



Dr. Neeraj
Nischal


Dr. Neeraj
Nischal
BAITUA


Dr. Naval K.
Vikram


Dr. Piyush
Ranjan


Dr. Navreet
Wig


Prof. R. Soel


Dr. Ashutosh
Biswas