

### QBC System:

1. It should come with an option to be attached with any compound microscope.
2. It should be easily attachable as an objective with a detachable filter arm.
3. Objective lenses should be 60X (Oil) for fluorescent microscopy of QBC tubes/capillaries.
4. If 100X (oil) objective is available for fluorescent microscopy of QBC tubes, its price should also be quoted alongwith.
5. The Light source should be LED based
6. The light source should have an average life of 20,000 hrs
7. Should have multiple power options
8. It should have multiple Filters Such as :
  - Excitation filter 385-480nm
  - Dichoric filter 510 nm
  - Emission Filter 480 nm
9. It should have a light output at 410-511nm
10. It should have a Excitation output at 410-480nm
11. A compatible voltage stabilizer should be provided
12. Should provide a comprehensive warranty for 5 years with 5 years **CMC** on the QBC system
13. User list must be submitted along with offer.
14. Company should be ISO/ISI/EU(European Union) certification.

**It will be purchased as a proprietary item**

Prepared by



Dr. Nishant Verma

Checked by



Dr. Hitender Gautam

*alok*



Corporate Address:  
Drucker Diagnostics  
168 Bradford Drive  
Port Matilda, PA 16870  
Phone: (814)-692-7661  
Fax: (814)-692-7662

Manufacturing Address:  
Drucker Diagnostics  
200 Shady Lane, Suite 170  
Phillipsburg, PA 16866  
Phone: (814)-342-6205  
Fax: (814)-342-4510



### Letter of Authorization

16 October 2017

**CLONEZYME BIOTEK**  
92, CHARUCHANDRA PLACE (EAST)  
GROUND FLOOR, TOLLYGUNE  
KOLKATA-700033, INDIA

This letter serves to announce and confirm the appointment of "CLONEZYME BIOTEK" as a distributor for Drucker Diagnostics.

This appointment is for the product families detailed in **Schedule 2** which are manufactured and sold by Drucker Diagnostics. These rights are appointed to "CLONEZYME BIOTEK" for countries enumerated in **Schedule 1**.

This authorization is valid from this date and will be maintained until December 31, 2019, unless it is specifically revoked or terminated. Drucker Diagnostics reserves the right to terminate this agreement at any time at its sole discretion. This authorization supersedes and replaces any agreement signed before this date.

The company "CLONEZYME BIOTEK" is fully authorized and has our full support to prepare letters, documents, illustrative technical pages, flyers, brochures and manuals in the appropriate language needed, and confirm authenticity of commercial literature and product catalogs and/or prepare the respective translations as may be requested or needed for technical and economical quotations to be presented in Public Tenders, Restrictive Invitation Tenders or directly to Private customers, Government Agencies and Universities.

The company "CLONEZYME BIOTEK" is fully authorized and has our support to participate in all kinds of commercial activities related and required to promote, quote and sell our products in Public Tenders, Restricted Invitation Tenders or directly to Private customers, Government Agencies, and Universities.


The company "CLONEZYME BIOTEK" is fully authorized and has our support to provide Technical Service, Customer Training, Organize Seminars and Participate in Local Congresses as required to maintain the Customer Satisfaction with our Product Line.

The company "CLONEZYME BIOTEK" is responsible to fully comply with U.S. laws and export regulations and to refrain from selling to any prohibited persons, entities, or end users as defined by the U.S. government.

Stephen Naspinski  
Quality Assurance Manager  
Date: 16 October 2017



Commonwealth of Pennsylvania.
County of <u>Centre</u>
Affirmed by <u>Tracey Bone</u>
On this <u>16th</u> day of <u>Oct</u> , <u>2017</u>





Corporate Address:  
Drucker Diagnostics  
168 Bradford Drive  
Port Matilda, PA 16870  
Phone: (814)-692-7661  
Fax: (814)-692-7662

Manufacturing Address:  
Drucker Diagnostics  
200 Shady Lane, Suite 170  
Philipsburg, PA 16866  
Phone: (814)-342-6205  
Fax: (814)-342-4510



**Drucker Diagnostics**  
Innovative Solutions for a Healthier World

## SCHEDULE 1

The products detailed in **Schedule 2** are authorized for sale non-exclusively by "CLONEZYME BIOTEK" in the following countries:

- India
- Afghanistan

## SCHEDULE 2

The following products are included as the products authorized for distribution by "CLONEZYME BIOTEK" when sold in the regions detailed in **Schedule 1**:

- QBC Paralens Advance Product Family
- QBC F.A.S.T. AFB Product Family
- QBC F.A.S.T. Malaria Product Family
- QBC Autoread Product Family
- QBC AccuTubes
- QBC Venous Tubes
- QBC Capillary Tubes
- QBC Malaria Tubes
- QBC STAR Product Family
- QBC STAR Tubes

QBC Diagnostics  
168 Bradford Drive  
Port Matilda, PA 16870  
Phone: (814)-692-7661  
Fax: (814)-692-7662



---

### PROPRIETARY CERTIFICATE

Date : 01-March-2016.

This is to certify that we, QBC Diagnostics, 168 Bradford Drive, Port Matilda, PA 16870, USA are sole Manufacturers (OEM) of following products .

- QBC Paralens Advance Product Family
- QBC F.A.S.T. AFB Product Family
- QBC F.A.S.T. Malaria Product Family
- QBC Autoread Product Family
- QBC AccuTubes
- QBC Venous Tubes
- QBC Capillary Tubes
- QBC Malaria Tubes
- QBC STAR Product Family
- QBC STAR Tubes

Michael Weaver  
Controller





# Monarch Scientific

*Distributors & Whole Sellers : Medical Equipment, Pharmaceuticals,  
Diagnostics & Surgical Items*

Telefax : 91-11-25682824  
Mobile : 91- 9871349497  
monarchscientific@yahoo.com  
IV/1/52, First Floor,  
Gopi Nath Bazar,  
Delhi Cantt – 110010  
GST No. –07AAJFM3375R1Z1

MQ/3897

21st May, 2018

To,

The HEAD,  
Department of Microbiology,  
AIIMS  
Ansari Nagar  
NEW DELHI- 110029

## QUOTATION

SL.	Cat_No.	Description	Qty.	Unit Price ( INR )	Price ( INR )
1	424331	QBC ParaLens Advance 60x & 100x, with QBC FM Scan – Micro Camera with Software	01	6,98,500.00	6,98,500.00
		Manufactured By : Drucker Dignostics / QBC Diagnostics, USA.			
		TOTAL : ( INR )			6,98,500.00
		ADD: GST@18% : (INR)			1,25,730.00
		GRAND TOTAL (INR) :			8,24,230.00
TOTAL (In Words): INR EIGHT LAKH TWENTY FOUR THOUSAND TWO HUNDRED THIRTY ONLY					

### TERMS & CONDITIONS

1. DELIVERY: 25-30 days from the date of order.
2. PAYMENT: 100% after delivery of the equipment  
In Favour of : "Monarch Scientific"
3. VALIDITY: 90Days.
4. WARRANTY: 5 (Five) Years from the date of installation.
5. Principal authorization letter is attached.

For **MONARCH SCIENTIFIC**  
( AUTHORISED SIGNATORY )



Celebrating 16 years of Excellence .... 2002 - 2018







# Monarch Scientific

*Distributors & Whole Sellers : Medical Equipment, Pharmaceuticals,  
Diagnostics & Surgical Items*

Telefax : 91-11-25682824  
Mobile : 91- 9871349497  
monarchscientific@yahoo.com  
IV/1/52, First Floor,  
Gopi Nath Bazar,  
Delhi Cantt - 110010  
GST No. -07AAJFM3375R1ZI

MQ/3897

21st May, 2018

To,  
The HEAD,  
Department of Microbiology,  
AIIMS  
Ansari Nagar  
NEW DELHI- 110029

## QUOTATION

SL.	Description	Qty.	Unit Price ( INR )	Price ( INR )
1.	Computer with UPS	01	FOC	FOC
2.	Malaria Kit	01	FOC	FOC
3.	AFB / Slide	01	FOC	FOC

### TERMS & CONDITIONS

- DELIVERY:** 25-30 Days from the date of order.
- PAYMENT:** 100% after delivery of the equipment  
In Favour of : "Monarch Scientific"
- VALIDITY:** 90 Days.
- WARRANTY:** 5 (Five) Years from the date of installation.
- Above mentioned goods will be supplied along with QBC Paralens advanced as FOC.

For **MONARCH SCIENTIFIC**  
( AUTHORISED SIGNATORY )



Celebrating 16 years of Excellence .... 2002 - 2018





# CloneZyme Biotech

REF: CB/KOL/AL-AIIMS/026/18-19

**Dated:** 17-April-2018

**Tender Ref No. :**

**Date :**

The Director  
AIIMS  
Ansari Nagar,  
NEW DELHI - 110029.

## SUBJECT :-AUTHORISATION LETTER

This is to certify that **M/s MONARCH SCIENTIFIC** having below mentioned address is our authorised distributor:-

**M/s MONARCH SCIENTIFIC**  
**IV/1/52 (1<sup>st</sup> Floor),**  
**Gopi Nath Bazaar,**  
**Delhi Cantt- 110010**

The above mentioned party is authorised to quote & supply the quoted materials as per your valuable order, raise their bills and collect payments thereof till **31<sup>st</sup> March, 2019**.

Thanking you and assuring you of our best services and quality products at all times.

For **Clonezyme Biotech**



[ Warinder Singh ]

---

**Kolkata Office :** 92, Charu Chandra Place (East), Kolkata 700033, West Bengal, INDIA.

**Tel :** +91 33 40616274. **Mob :** +91 98315-62989. **Email:** [clonezyme.biotech@gmail.com](mailto:clonezyme.biotech@gmail.com).

# CloneZyme Biotech

Quotation Ref. No. : CB/KOL/CMSD/0246/18-19

Date : 05-May-2018

The HOD, Dept of Microbiology,  
**All India Institute of Medical Sciences**  
New Delhi.

Dear Sir,

This is to certify that we have quoted our lowest price for QBC ParaLens advance System for Malaria, manufactured by **Drucker Diagnostics USA**. We are also confirming that we have not quoted a price lower than this to any Government or other Institute.

Thanking you,

For **Clonezyme Biotech**

A handwritten signature in black ink is written over a circular stamp. The stamp contains the text "CLONEZYME BIOTEK" around the perimeter and a star in the center.

[ Authorized Signatory ]

---

**Kolkata Office :** 92, Charu Chandra Place (East), Kolkata 700033, West Bengal, INDIA.


**Tel :** +91 33 40616274. **Mob :** +91 98315-62989. **Email:** [clonezyme.biotek@gmail.com](mailto:clonezyme.biotek@gmail.com).




ALL INDIA INSTITUTE OF MEDICAL SCIENCES  
ANSARI NAGAR, NEW DELHI - 110 029

PROPRIETORY/SPECIFIC BRAND GOODS CERTIFICATE

1. Item/type/Model No. required **QBC PARALENS ADVANCE**  
along with specifications.
2. Is the item a spare part attachment  
or accessory for an existing equipment. **NA**
3. Name of Manufacturers supplier of the  
item proposed by the indenter. **M/S DRUCKER DIAGNOSTICS**
4. Are they sole manufacturers/sole distri-  
butors of the item. **YES**
5. Is there any other item with similar/  
equivalent specification available in the  
market to meet the job requirement envisaged.  
If the answer is yes, why the same can't be  
procured. Demanding Officer should bring out  
comparative functional advantages/cost  
effectiveness of the recommended item from  
those offered by others. **NA**
6. What were the efforts made to locate alternative  
source of supply or use other substitutes. **NA**
7. Why open/limited tender can't be resorted to, for  
locating alternative source. **PROPRIETARY ITEM**
8. Are the proprietary items certifying that the rates  
are reasonable or not. **REASONABLE**
9. Any other justifications for procuring item from  
single source. **PROPRIETARY**

  
Signature of indenter  
(Demanding Officer)



  
COUNTERSIGNED  
(Head of the Department)

**Dr. GITA SATPATHY**  
Prof. & Head  
Department of Microbiology  
All India Institute of Medical Sciences  
Ansari Nagar, New Delhi-110029