

ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI-29.
STORES SECTION (DO)


Ref. No. 26/Stores(DO)/Uro./PAC/2019-20/FSC Dated-21/11/2019

Sub:- Purchase of "Intra Corporeal Pneumatic Lithotripter – 01 No." for the Department of Urology at AIIMS, New Delhi-110029, on proprietary basis inviting comments thereon.

The Institute is in the process to purchase "Intra Corporeal Pneumatic Lithotripter – 01 No. at AIIMS, New Delhi from M/s. Electro Medical Systems, Switzerland through M/s. Healthware Pvt. Ltd. The PAC Certifications by M/s. Electro Medical Systems, Switzerland as well as the user department are attached.

The above documents are being uploaded for open information to submit objections, comments, if any, from any manufacturer regarding proprietary nature of the equipment/item within 15 day from the date of issue/uploading of the notification giving reference No. 26/Stores(DO)/Uro./PAC/2019-20/FSC. The comments should be received in office of Stores Officer (FSC), Store Section (DO), Animal House Building, Near Biotechnology Building at AIIMS on or before 12/12/2019 upto 12.30 p.m. failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

Yours faithfully,


SR. STORES OFFICER (DO)

Encl: Related documents enclosed.

**ALL INDIA INSTITUTE OF MEDICAL SCIENCES
ANSARI NAGAR, NEW DELHI – 110 029**

(7)

Proprietary Article Certificate (PAC) (Machinery & Equipment)

1.	The indented goods are manufactured by M/s	:	E.M.S. Electro Medical Systems,S.A.
2.	Item Name	:	INTRA CORPOREAL PNEUMATIC LITHOTRIPTER
3.	Model No.	:	EMS SWISS LITHOCLAST-II SYSTEM
4.	Vital Technical Performance Parameters required which makes the requirement Proprietary##	:	Swiss Lithoclast –II with proprietary technologies namely, an autoclavable and completely sealed hand piece to prevent contamination and a manual over, ride feature to complete a procedure in the event of the failure of the foot pedal.
5.	No other make or model is acceptable for the following reasons	:	Only Single Manufacturer

It is certified that market survey has been done and found that no other manufacturer is manufacturing similar / equipment specifications which can fulfill the vital requirements of end user.

##Note: 1.: TSEC should clearly mention the vital functional parameters requirements which end user essentially require and are manufactured by only one manufacturer mentioned in serial no.1.

DR. ANUP KUMAR
MCh (Urology) AIIMS
DNB (Urology) MNAMS
Uro Oncology Fellow in Robotics (USA)
Central Health Services
Professor and Head of Department
Department of Urology & Renal Transplant
Nodal Officer for Cancer & Renal Transplant
VMMC & Safdarjung Hospital, New Delhi

Dr. RAJEEV KUMAR
आचार्य / Professor
मूत्ररोग विज्ञान विभाग / Deptt. of Urology
अ.भा.आ.सं., नई दिल्ली-29/A.I.I.M.S., New Delhi-29

Dr. PRADEEP SINGH
आचार्य / Associate Professor
मूत्ररोग विज्ञान विभाग / Department of Urology
अ.भा.आ.सं., नई दिल्ली / A.I.I.M.S., New Delhi-29

Dr. SANJAY
सहायक आचार्य / Assistant Professor
मूत्ररोग विज्ञान विभाग / Department of Urology
अ.भा.आ.सं., नई दिल्ली / A.I.I.M.S., New Delhi-29

Dr. ANISH KUMAR
आचार्य / Professor
मूत्ररोग विज्ञान विभाग / Department of Urology
अ.भा.आ.सं., नई दिल्ली / A.I.I.M.S., New Delhi-29

Dr. MANOJ K.
सहायक आचार्य / Assistant Professor
मूत्ररोग विज्ञान विभाग / Department of Urology
अ.भा.आ.सं., नई दिल्ली / A.I.I.M.S., New Delhi-29

Dr. BRULASHANI NAYAK
सहायक आचार्य / Associate Professor
मूत्ररोग विज्ञान विभाग / Department of Urology
अ.भा.आ.सं., नई दिल्ली / A.I.I.M.S., New Delhi-29

Dr. SIDDHARTH JA
सहायक आचार्य / Assistant Professor
मूत्ररोग विज्ञान विभाग / Department of Urology
अ.भा.आ.सं., नई दिल्ली / A.I.I.M.S., New Delhi-29

SPECIFICATIONS FOR INTRA CORPOREAL PNEUMATIC LITHOTRIPTER

An Intra corporeal Lithotripter should have following specifications:

- It should be light weight, compact and easy to operate.
- Repetition rate from 1 to 12 Hz.
- Ballistic energy to fragment even the hardest stones.
- A dual pedal footswitch to operate.
- Single or multiple mode selection should be from the dual pedal foot switch itself.
- The hand piece should be sterilizable.
- Facility to integrate the hand piece to the suction equipment should be provided.
- Dedicated medical grade compressor should be supplied
- Should be able to withstand pressures from 3.5 to 5 Bar
- Displacement measured under standard conditions (2 bar) should be ideally less than 1.5 mm
- Facility to collect the stone fragments.
- Manual override to operate the machine in case of electronic failure of the control unit should be available.
- The unit should be supplied with movable trolley manufactured by the same manufacturer
- The unit should be quoted with suitable UPS of adequate rating for 30 minutes power backup.

PNEUMATIC PROBES:

- a) 0.8 mm diameter pneumatic probe- qty 10 numbers.
- b) 1 mm diameter pneumatic probe - qty 10 numbers.
- c) 1.6 mm diameter pneumatic probe - qty 5 numbers.
- d) 2 mm diameter pneumatic probe - qty 5 numbers.
- e) 1.3 mm diameter for use with mini perc – 2 numbers.
- f) 0.8 mm diameter for use with mini perc – 2 numbers.
- g) 0.89 mm diameter, 900-1000mm length for use with Flexible Ureteroscope – 3 numbers.

[Handwritten signatures and initials at the bottom of the page, including "Siddhant", "Rhe", "Mangal", and others.]

- h) 0.89 mm diameter, 590-610mm length for use with Flexible Nephroscope – 3 numbers.

SUCTION PROBES:

- a) Compatible Suction Probe with 2 mm pneumatic probe – 2 numbers
- b) Compatible Suction Probe with 1.6 mm pneumatic probe – 2 numbers
- c) Compatible Suction Probe with 0.08 mm pneumatic probe – 2 numbers
- d) 0.8 mm pneumatic probe to be used with suction facility- 2 numbers
- e) 1.6 mm pneumatic probe to be used with suction facility – 2 numbers

TERM AND CONDITIONS

- The unit should be supplied with minimum warranty period of 60 months of satisfactory installation.
- The warranty must cover each and every part of the unit.
- The unit should meet the international safety standard for human use (CE and USFDA).
- The supplier should quote the CMC for 5 years beyond the warranty.
- CMC must cover all the items of the main equipment.
- The bidder must quote the price of all the items of essential accessories separately.
- The price of all the accessories will be frozen during the period of warranty and maintenance.

Handwritten signatures and initials are present at the bottom of the page, including a large signature on the left, a signature with 'S. Sankar' above it, a signature with 'Roh' above it, a signature with 'Mang' above it, and several other initials and signatures.

866.19/afe

Nyon, September 19th, 2019

MANUFACTURER'S AUTHORISATION FORM

The Director
All India Institute of Medical Sciences,
Ansari Nagar, New Delhi – 110029, India.

Dear Sir,

We, E.M.S. Electro Medical Systems S.A. who are proven and reputable manufacturers of **Intra Corporeal Pneumatic Lithotripter** having factories at Chemin de La Vuarpilliére 31, CH-1260 Nyon, Switzerland, hereby authorise Messrs **Healthware Private Limited**, Serene Towers, 4th Floor, 8-2-623/A, Road No. 10, Banjara Hills, Hyderabad – 500 034 to submit a bid, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above goods manufactured by us.


We further confirm that no supplier or firm or individual other than Messrs. **M/s Healthware Private Limited** is authorised to submit a bid, process the same further and enter into a contract with you against your requirement as contained in the above referred TE documents for the above good manufactured by us.

We also hereby extend our full warranty, CAMC as applicable as per clause 15 of the General Conditions of Contract, read with modification, if any, in the Special Conditions of Contract for the goods and services offered for supply by the above firm against this TE document.

We also hereby confirm that we would be responsible for the satisfactory execution of contract placed on the authorized agent and the spares for the equipment shall be available for at least 10 years from the date of supply of equipment.

We also confirm that the price quoted by our agent shall not exceed the price which we would have quoted directly.

E.M.S. Electro Medical Systems S.A.


Alain Siegrist
CFO

EMS[®]
ELECTRO MEDICAL SYSTEMS S.A.
Ch. de la Vuarpilliére 31
CH-1260 NYON (Switzerland)
Tel. +41 22 994 47 00
Fax +41 22 994 47 01


Andrea Blumer
Senior Legal Advisor



865.19/afe

Nyon, September 19th, 2019.

Head of Department
Department of Urology,
All India Institute of Medical Sciences
New Delhi

Sub: Proprietary Certificate for Swiss Lithoclast II

Dear Sir,

We, **M/S E.M.S. ELECTRO MEDICAL SYSTEMS S.A.**, having our registered office at ch. de la Vuarpillière 31, CH-1260 Nyon, Switzerland, certify & confirm that EMS Swiss Lithoclast II is being solely manufactured by us and is proprietary item of EMS.


This is to certify that the Swiss Lithoclast II, with proprietary technologies namely, an autoclavable and completely sealed hand piece to prevent contamination and a manual over ride feature to complete a procedure in the event of the failure of the foot pedal, is being manufactured by us.


This certificate is valid from date of signature until March 31st, 2020.

E.M.S. Electro Medical Systems S.A.


Alain Siegrist
CFO

EMS⁺
ELECTRO MEDICAL SYSTEMS SA
Ch. de la Vuarpillière 31
CH-1260 NYON (Switzerland)
Tél. +41 22 994 47 00
Fax +41 22 994 47 01


Andrea Blumer
Senior Legal Advisor


Dr. RAJEEV KUMAR
Professor
Urology / Dept. of Urology
All India Institute of Medical Sciences, New Delhi-29

